

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050523\
 Data File : VX035533.D
 Acq On : 05 May 2023 14:05
 Operator : JC/MD
 Sample : 02615-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 003-1

Quant Time: May 08 00:50:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	289123	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	486428	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	445344	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	209134	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	176325	50.550	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.100%
35) Dibromofluoromethane	5.385	113	148551	49.757	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.520%
50) Toluene-d8	8.647	98	572273	50.812	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.620%
62) 4-Bromofluorobenzene	11.079	95	219635	51.841	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.680%
Target Compounds						
						Qvalue
16) Acetone	2.374	43	4074	2.689	ug/l	90
29) Tetrahydrofuran	5.038	42	1053	0.743	ug/l #	84
30) Chloroform	5.099	83	24676	4.438	ug/l	86
64) Tetrachloroethene	9.275	164	1413	0.414	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050523\
Data File : VX035533.D
Acq On : 05 May 2023 14:05
Operator : JC/MD
Sample : 02615-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
003-1

Quant Time: May 08 00:50:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

