

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050523\
 Data File : VX035535.D
 Acq On : 05 May 2023 14:52
 Operator : JC/MD
 Sample : 02615-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 002-2

Quant Time: May 08 00:51:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|--------|----------|
| ----- | | | | | | |
| Internal Standards | | | | | | |
| 1) Pentafluorobenzene | 5.550 | 168 | 268454 | 50.000 | ug/l | 0.00 |
| 34) 1,4-Difluorobenzene | 6.763 | 114 | 458239 | 50.000 | ug/l | 0.00 |
| 63) Chlorobenzene-d5 | 10.055 | 117 | 418608 | 50.000 | ug/l | 0.00 |
| 72) 1,4-Dichlorobenzene-d4 | 12.024 | 152 | 187450 | 50.000 | ug/l | 0.00 |
| System Monitoring Compounds | | | | | | |
| 33) 1,2-Dichloroethane-d4 | 5.958 | 65 | 166641 | 51.452 | ug/l | 0.00 |
| Spiked Amount | 50.000 | Range | 78 - 117 | Recovery | = | 102.900% |
| 35) Dibromofluoromethane | 5.385 | 113 | 141793 | 50.415 | ug/l | 0.00 |
| Spiked Amount | 50.000 | Range | 75 - 124 | Recovery | = | 100.820% |
| 50) Toluene-d8 | 8.647 | 98 | 539044 | 50.806 | ug/l | 0.00 |
| Spiked Amount | 50.000 | Range | 92 - 112 | Recovery | = | 101.620% |
| 62) 4-Bromofluorobenzene | 11.079 | 95 | 199249 | 49.922 | ug/l | 0.00 |
| Spiked Amount | 50.000 | Range | 83 - 123 | Recovery | = | 99.840% |
| Target Compounds | | | | | | |
| | | | | | | Qvalue |
| 16) Acetone | 2.379 | 43 | 4321 | 3.072 | ug/l | 98 |
| 30) Chloroform | 5.098 | 83 | 57533 | 11.143 | ug/l | 96 |
| 47) Bromodichloromethane | 7.824 | 83 | 29007 | 7.218 | ug/l # | 99 |
| 60) Dibromochloromethane | 9.525 | 129 | 11384 | 3.732 | ug/l | 100 |
| ----- | | | | | | |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050523\
 Data File : VX035535.D
 Acq On : 05 May 2023 14:52
 Operator : JC/MD
 Sample : 02615-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 002-2

Quant Time: May 08 00:51:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

