

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050523\
 Data File : VX035536.D
 Acq On : 05 May 2023 15:15
 Operator : JC/MD
 Sample : 02615-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 003-2

Quant Time: May 08 00:51:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	265452	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	445690	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	397630	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	179649	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	161057	50.291	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.580%	
35) Dibromofluoromethane	5.385	113	137903	50.412	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.820%	
50) Toluene-d8	8.646	98	516315	50.034	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.060%	
62) 4-Bromofluorobenzene	11.079	95	192664	49.631	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 99.260%	
Target Compounds						
						Qvalue
30) Chloroform	5.092	83	18917	3.705	ug/l	92
44) Trichloroethene	7.128	130	971	0.288	ug/l	54
47) Bromodichloromethane	7.823	83	970	0.248	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050523\
Data File : VX035536.D
Acq On : 05 May 2023 15:15
Operator : JC/MD
Sample : 02615-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
003-2

Quant Time: May 08 00:51:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

