

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050523\
 Data File : VX035541.D
 Acq On : 05 May 2023 17:11
 Operator : JC/MD
 Sample : PB152615TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB152615TB

Quant Time: May 08 00:52:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	282764	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	477477	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	435745	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	207119	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	174775	51.233	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 102.460%		
35) Dibromofluoromethane	5.385	113	147615	50.370	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 100.740%		
50) Toluene-d8	8.647	98	563021	50.928	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 101.860%		
62) 4-Bromofluorobenzene	11.079	95	217271	52.244	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 104.480%		
Target Compounds						
						Qvalue
11) Tert butyl alcohol	2.983	59	4540	7.309	ug/l #	90
16) Acetone	2.386	43	32520	21.949	ug/l	99
20) Methylene Chloride	2.788	84	12139	3.573	ug/l	93
25) 2-Butanone	4.568	43	23517	10.875	ug/l	100
37) Ethyl Acetate	4.714	43	58319	13.954	ug/l #	95
43) Isopropyl Acetate	6.336	43	226268	32.219	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050523\
Data File : VX035541.D
Acq On : 05 May 2023 17:11
Operator : JC/MD
Sample : PB152615TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB152615TB

Quant Time: May 08 00:52:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

