

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050523\
 Data File : VX035544.D
 Acq On : 05 May 2023 18:21
 Operator : JC/MD
 Sample : 02598-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ETGI-328

Quant Time: May 08 00:52:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	276689	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	462140	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	430563	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	206201	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	170263	51.006	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 102.020%		
35) Dibromofluoromethane	5.385	113	144547	50.960	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 101.920%		
50) Toluene-d8	8.647	98	544394	50.877	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 101.760%		
62) 4-Bromofluorobenzene	11.079	95	212389	52.765	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 105.520%		
Target Compounds						
						Qvalue
11) Tert butyl alcohol	2.977	59	4797	7.892	ug/l	# 88
16) Acetone	2.386	43	35911	24.770	ug/l	97
20) Methylene Chloride	2.794	84	14161	4.260	ug/l	98
25) 2-Butanone	4.568	43	24721	11.682	ug/l	95
37) Ethyl Acetate	4.715	43	27278	6.743	ug/l	97
43) Isopropyl Acetate	6.336	43	121508	17.876	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050523\
Data File : VX035544.D
Acq On : 05 May 2023 18:21
Operator : JC/MD
Sample : 02598-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ETGI-328

Quant Time: May 08 00:52:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

