

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050523\
 Data File : VX035545.D
 Acq On : 05 May 2023 18:44
 Operator : JC/MD
 Sample : 02643-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-PRO-07-90-05032023

Quant Time: May 08 00:53:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.555	168	263009	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	436535	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	405442	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	196658	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	161186	50.798	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.600%
35) Dibromofluoromethane	5.391	113	136271	50.860	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.720%
50) Toluene-d8	8.646	98	515156	50.969	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.940%
62) 4-Bromofluorobenzene	11.079	95	200602	52.760	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.520%
Target Compounds						Qvalue
44) Trichloroethene	7.128	130	523975	158.485	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050523\
Data File : VX035545.D
Acq On : 05 May 2023 18:44
Operator : JC/MD
Sample : 02643-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-PRO-07-90-05032023

Quant Time: May 08 00:53:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

