

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050523\
 Data File : VX035546.D
 Acq On : 05 May 2023 19:08
 Operator : JC/MD
 Sample : 02643-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-PRO-06-99-05042023

Quant Time: May 08 00:53:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	277409	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	469449	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	442425	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	211608	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	175074	52.311	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.620%
35) Dibromofluoromethane	5.385	113	147286	51.117	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.240%
50) Toluene-d8	8.647	98	552232	50.806	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.620%
62) 4-Bromofluorobenzene	11.079	95	216898	53.046	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.100%
Target Compounds						
						Qvalue
12) 1,1-Dichloroethene	2.319	96	40975	15.208	ug/l	98
16) Acetone	2.386	43	3149	2.166	ug/l	95
27) cis-1,2-Dichloroethene	4.495	96	6525	1.897	ug/l	88
44) Trichloroethene	7.129	130	1896414	533.385	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050523\
Data File : VX035546.D
Acq On : 05 May 2023 19:08
Operator : JC/MD
Sample : 02643-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-PRO-06-99-05042023

Quant Time: May 08 00:53:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

