

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX050619\
 Data File : VX009348.D
 Acq On : 06 May 2019 14:18
 Operator : JC/SP
 Sample : K2707-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-MW201D1-GW-20190502

Quant Time: May 07 07:36:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 04 07:18:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	189432	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	303223	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	263462	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	110352	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	123871	47.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.80%	
35) Dibromofluoromethane	5.50	113	92544	48.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.10%	
50) Toluene-d8	8.71	98	376478	48.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.84%	
62) 4-Bromofluorobenzene	11.13	95	123839	44.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.58%	
Target Compounds						
44) Trichloroethene	7.21	130	8001	3.795	ug/l	97
64) Tetrachloroethene	9.34	164	1567	0.867	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050619\
Data File : VX009348.D
Acq On : 06 May 2019 14:18
Operator : JC/SP
Sample : K2707-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-MW201D1-GW-20190502

Quant Time: May 07 07:36:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Sat May 04 07:18:36 2019
Response via : Initial Calibration

