

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX050619\
 Data File : VX009349.D
 Acq On : 06 May 2019 14:41
 Operator : JC/SP
 Sample : K2663-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ST007MW053-190429

Quant Time: May 07 07:39:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 04 07:18:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	189017	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	306042	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	268023	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	115862	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	125108	48.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.98%	
35) Dibromofluoromethane	5.50	113	93376	48.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.06%	
50) Toluene-d8	8.71	98	381144	49.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.14%	
62) 4-Bromofluorobenzene	11.13	95	125984	44.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.28%	
Target Compounds						
4) Vinyl Chloride	1.40	62	3392	1.463	ug/l #	80
16) Acetone	2.45	43	3631	2.621	ug/l	96
27) cis-1,2-Dichloroethene	4.60	96	4011	1.747	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050619\
Data File : VX009349.D
Acq On : 06 May 2019 14:41
Operator : JC/SP
Sample : K2663-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ST007MW053-190429

Quant Time: May 07 07:39:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Sat May 04 07:18:36 2019
Response via : Initial Calibration

