

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX050619\
 Data File : VX009362.D
 Acq On : 06 May 2019 19:46
 Operator : JC/SP
 Sample : K2675-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OUTFALL001-WK-1

Quant Time: May 07 11:12:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 04 07:18:36 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	185275	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	299047	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	261179	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	112337	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	122015	48.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.48%	
35) Dibromofluoromethane	5.50	113	91329	48.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.16%	
50) Toluene-d8	8.71	98	374770	49.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.76%	
62) 4-Bromofluorobenzene	11.13	95	123846	44.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.82%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	24646	10.952	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050619\
Data File : VX009362.D
Acq On : 06 May 2019 19:46
Operator : JC/SP
Sample : K2675-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OUTFALL001-WK-1

Quant Time: May 07 11:12:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Sat May 04 07:18:36 2019
Response via : Initial Calibration

