

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX050619\
 Data File : VX009365.D
 Acq On : 06 May 2019 20:56
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 07 08:21:21 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 04 07:18:36 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	83	0.00
2 T	Dichlorodifluoromethane	0.461	0.407	11.7	75	0.00
3 P	Chloromethane	0.629	0.587	6.7	80	0.00
4 C	Vinyl Chloride	0.613	0.575	6.2#	79	0.00
5 T	Bromomethane	0.368	0.361	1.9	87	0.00
6 T	Chloroethane	0.394	0.403	-2.3	88	0.00
7 T	Trichlorofluoromethane	0.746	0.720	3.5	82	0.00
8 T	Diethyl Ether	0.353	0.350	0.8	86	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.468	0.441	5.8	80	0.00
10 T	Methyl Iodide	0.483	0.470	2.7	77	0.00
11 T	Tert butyl alcohol	0.158	0.180	-13.9	97	0.00
12 CM	1,1-Dichloroethene	0.469	0.456	2.8#	81	0.00
13 T	Acrolein	0.143	0.128	10.5	77	0.00
14 T	Allyl chloride	1.143	1.087	4.9	82	0.00
15 T	Acrylonitrile	0.386	0.409	-6.0	92	0.00
16 T	Acetone	0.367	0.356	3.0	84	0.00
17 T	Carbon Disulfide	1.317	1.166	11.5	73	0.00
18 T	Methyl Acetate	0.872	0.937	-7.5	95	0.00
19 T	Methyl tert-butyl Ether	1.822	1.860	-2.1	87	0.00
20 T	Methylene Chloride	0.593	0.562	5.2	83	0.00
21 T	trans-1,2-Dichloroethene	0.509	0.481	5.5	80	0.00
22 T	Diisopropyl ether	2.199	2.206	-0.3	85	0.00
23 T	Vinyl Acetate	1.883	1.960	-4.1	86	0.00
24 P	1,1-Dichloroethane	1.085	1.059	2.4	84	0.00
25 T	2-Butanone	0.570	0.612	-7.4	91	0.00
26 T	2,2-Dichloropropane	0.807	0.672	16.7	70	0.00
27 T	cis-1,2-Dichloroethene	0.607	0.600	1.2	85	0.00
28 T	Bromochloromethane	0.487	0.415	14.8	70	0.00
29 T	Tetrahydrofuran	0.366	0.402	-9.8	92	0.00
30 C	Chloroform	0.981	0.993	-1.2#	87	0.00
31 T	Cyclohexane	1.009	0.935	7.3	77	0.00
32 T	1,1,1-Trichloroethane	0.817	0.804	1.6	84	0.00
33 S	1,2-Dichloroethane-d4	0.683	0.681	0.3	82	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	85	0.00
35 S	Dibromofluoromethane	0.314	0.312	0.6	83	0.00
36 T	1,1-Dichloropropene	0.465	0.432	7.1	81	0.00
37 T	Ethyl Acetate	0.647	0.685	-5.9	92	0.00
38 T	Carbon Tetrachloride	0.422	0.406	3.8	83	0.00
39 T	Methylcyclohexane	0.571	0.512	10.3	77	0.00
40 TM	Benzene	1.438	1.376	4.3	83	0.00
41 T	Methacrylonitrile	0.385	0.390	-1.3	90	0.00
42 TM	1,2-Dichloroethane	0.516	0.506	1.9	87	0.00
43 T	Isopropyl Acetate	1.036	1.069	-3.2	89	0.00
44 TM	Trichloroethene	0.348	0.326	6.3	83	0.00
45 C	1,2-Dichloropropane	0.407	0.399	2.0#	85	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX050619\
 Data File : VX009365.D
 Acq On : 06 May 2019 20:56
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 07 08:21:21 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 04 07:18:36 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.236	0.230	2.5	86	0.00
47 T	Bromodichloromethane	0.472	0.468	0.8	86	0.00
48 T	Methyl methacrylate	0.514	0.533	-3.7	89	0.00
49 T	1,4-Dioxane	0.010	0.011	-10.0	96	0.00
50 S	Toluene-d8	1.269	1.262	0.6	82	0.00
51 T	4-Methyl-2-Pentanone	0.677	0.732	-8.1	93	0.00
52 CM	Toluene	0.874	0.840	3.9#	83	0.00
53 T	t-1,3-Dichloropropene	0.533	0.534	-0.2	84	0.00
54 T	cis-1,3-Dichloropropene	0.589	0.580	1.5	82	0.00
55 T	1,1,2-Trichloroethane	0.357	0.355	0.6	86	0.00
56 T	Ethyl methacrylate	0.637	0.655	-2.8	87	0.00
57 T	1,3-Dichloropropane	0.628	0.618	1.6	86	0.00
58 T	2-Chloroethyl Vinyl ether	0.338	0.340	-0.6	83	0.00
59 T	2-Hexanone	0.532	0.576	-8.3	92	0.00
60 T	Dibromochloromethane	0.348	0.357	-2.6	87	0.00
61 T	1,2-Dibromoethane	0.360	0.360	0.0	86	0.00
62 S	4-Bromofluorobenzene	0.461	0.463	-0.4	83	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	85	0.00
64 T	Tetrachloroethene	0.343	0.331	3.5	85	0.00
65 PM	Chlorobenzene	1.000	0.980	2.0	85	0.00
66 T	1,1,1,2-Tetrachloroethane	0.363	0.365	-0.6	87	0.00
67 C	Ethyl Benzene	1.841	1.794	2.6#	84	0.00
68 T	m/p-Xylenes	0.673	0.660	1.9	83	0.00
69 T	o-Xylene	0.664	0.652	1.8	85	0.00
70 T	Styrene	1.154	1.160	-0.5	85	0.00
71 P	Bromoform	0.295	0.313	-6.1	90	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	0.00
73 T	Isopropylbenzene	3.636	3.468	4.6	84	0.00
74 T	N-amyl acetate	2.120	2.178	-2.7	88	0.00
75 P	1,1,2,2-Tetrachloroethane	1.355	1.362	-0.5	90	0.00
76 T	1,2,3-Trichloropropane	1.160	1.164	-0.3	89	0.00
77 T	Bromobenzene	0.902	0.877	2.8	87	0.00
78 T	n-propylbenzene	4.152	3.946	5.0	83	0.00
79 T	2-Chlorotoluene	2.562	2.400	6.3	86	0.00
80 T	1,3,5-Trimethylbenzene	3.061	2.935	4.1	85	0.00
81 T	trans-1,4-Dichloro-2-butene	0.453	0.441	2.6	83	0.00
82 T	4-Chlorotoluene	2.894	2.747	5.1	85	0.00
83 T	tert-Butylbenzene	2.978	2.856	4.1	86	0.00
84 T	1,2,4-Trimethylbenzene	3.073	2.966	3.5	85	0.00
85 T	sec-Butylbenzene	3.531	3.353	5.0	84	0.00
86 T	p-Isopropyltoluene	3.174	3.057	3.7	84	0.00
87 T	1,3-Dichlorobenzene	1.609	1.547	3.9	85	0.00
88 T	1,4-Dichlorobenzene	1.622	1.560	3.8	86	0.00
89 T	n-Butylbenzene	2.869	2.721	5.2	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX050619\
Data File : VX009365.D
Acq On : 06 May 2019 20:56
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: May 07 08:21:21 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Sat May 04 07:18:36 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.541	0.522	3.5	83	0.00
91 T	1,2-Dichlorobenzene	1.639	1.586	3.2	86	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.304	0.323	-6.3	91	0.00
93 T	1,2,4-Trichlorobenzene	1.121	1.115	0.5	86	0.00
94 T	Hexachlorobutadiene	0.572	0.539	5.8	83	0.00
95 T	Naphthalene	3.646	3.783	-3.8	88	0.00
96 T	1,2,3-Trichlorobenzene	1.136	1.138	-0.2	87	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6