


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX050619\
Data File : VX009337.D
Acq On : 06 May 2019 09:18
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Title : SW846 8260
Last Update : Tue May 07 08:24:47 2019



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1629

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.8	6799	PASS
75	95	30	60	51.0	17514	PASS
95	95	100	100	100.0	34338	PASS
96	95	5	9	6.7	2288	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	81.4	27952	PASS
175	174	5	9	7.8	2182	PASS
176	174	95	101	95.4	26677	PASS
177	176	5	9	6.7	1800	PASS