

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050622\
 Data File : VX028542.D
 Acq On : 06 May 2022 12:16
 Operator : JC/MD
 Sample : VX0506WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0506WBSD01

Quant Time: May 07 05:59:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	288852	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	451402	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	420997	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	223996	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	197154	52.347	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 104.700%	
35) Dibromofluoromethane	5.385	113	176320	55.698	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 111.400%	
50) Toluene-d8	8.653	98	610614	50.221	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 100.440%	
62) 4-Bromofluorobenzene	11.079	95	235704	50.701	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 101.400%	

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	42286	22.678	ug/l	99
3) Chloromethane	1.294	50	42106	21.000	ug/l	98
4) Vinyl Chloride	1.374	62	50512	17.346	ug/l	100
5) Bromomethane	1.605	94	60573	18.130	ug/l	97
6) Chloroethane	1.685	64	37845	15.953	ug/l	98
7) Trichlorofluoromethane	1.886	101	96257	14.990	ug/l	97
8) Diethyl Ether	2.136	74	37992	17.677	ug/l	96
9) 1,1,2-Trichlorotrifluo...	2.331	101	52630	22.500	ug/l	98
10) Methyl Iodide	2.453	142	56300	20.992	ug/l	99
11) Tert butyl alcohol	2.959	59	67876	87.655	ug/l	98
12) 1,1-Dichloroethene	2.319	96	45246	20.382	ug/l	96
13) Acrolein	2.233	56	57750	125.114	ug/l	98
14) Allyl chloride	2.666	41	66460	17.842	ug/l	92
15) Acrylonitrile	3.062	53	160601	104.637	ug/l	99
16) Acetone	2.380	43	129683	111.963	ug/l	93
17) Carbon Disulfide	2.514	76	91610	15.324	ug/l	99
18) Methyl Acetate	2.703	43	69593	20.782	ug/l	99
19) Methyl tert-butyl Ether	3.111	73	189588	20.481	ug/l	98
20) Methylene Chloride	2.788	84	58562	20.956	ug/l	98
21) trans-1,2-Dichloroethene	3.093	96	52584	17.418	ug/l	91
22) Diisopropyl ether	3.757	45	164816	19.427	ug/l #	81
23) Vinyl Acetate	3.721	43	705248	95.149	ug/l	100
24) 1,1-Dichloroethane	3.611	63	97086	19.380	ug/l	99
25) 2-Butanone	4.556	43	209839	94.075	ug/l	99
26) 2,2-Dichloropropane	4.477	77	77816	15.910	ug/l	99
27) cis-1,2-Dichloroethene	4.489	96	67425	18.730	ug/l	92
28) Bromochloromethane	4.903	49	48510	22.571	ug/l	91
29) Tetrahydrofuran	5.007	42	130639	89.824	ug/l	99
30) Chloroform	5.092	83	113778	18.912	ug/l	93
31) Cyclohexane	5.470	56	71166	15.509	ug/l	97
32) 1,1,1-Trichloroethane	5.385	97	95438	17.519	ug/l	99
36) 1,1-Dichloropropene	5.696	75	73455	17.789	ug/l	99
37) Ethyl Acetate	4.714	43	80193	19.139	ug/l	100
38) Carbon Tetrachloride	5.678	117	82807	17.764	ug/l	100
39) Methylcyclohexane	7.385	83	81734	15.687	ug/l	96
40) Benzene	6.043	78	226842	18.767	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050622\
 Data File : VX028542.D
 Acq On : 06 May 2022 12:16
 Operator : JC/MD
 Sample : VX0506WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0506WBSD01

Quant Time: May 07 05:59:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	42844	19.819	ug/l	99
42) 1,2-Dichloroethane	6.092	62	84644	18.865	ug/l	100
43) Isopropyl Acetate	6.342	43	127704	18.727	ug/l	98
44) Trichloroethene	7.129	130	66381	18.938	ug/l	96
45) 1,2-Dichloropropane	7.433	63	58949	19.924	ug/l	97
46) Dibromomethane	7.580	93	45996	19.818	ug/l	97
47) Bromodichloromethane	7.824	83	86012	19.282	ug/l	96
48) Methyl methacrylate	7.696	41	59951	19.149	ug/l	93
49) 1,4-Dioxane	7.665	88	31003	370.169	ug/l	94
51) 4-Methyl-2-Pentanone	8.574	43	427467	97.686	ug/l	97
52) Toluene	8.720	92	151878	18.086	ug/l	99
53) t-1,3-Dichloropropene	8.982	75	82748	17.455	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	92422	18.155	ug/l	94
55) 1,1,2-Trichloroethane	9.153	97	69077	20.246	ug/l	97
56) Ethyl methacrylate	9.116	69	95435	19.296	ug/l #	92
57) 1,3-Dichloropropane	9.311	76	109467	19.878	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.244	63	229627	98.953	ug/l	95
59) 2-Hexanone	9.433	43	328550	96.927	ug/l	95
60) Dibromochloromethane	9.525	129	71837	19.313	ug/l	97
61) 1,2-Dibromoethane	9.610	107	72561	19.385	ug/l	99
64) Tetrachloroethene	9.275	164	63261	19.238	ug/l	97
65) Chlorobenzene	10.079	112	175608	19.564	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.165	131	67163	20.086	ug/l	99
67) Ethyl Benzene	10.195	91	293181	19.479	ug/l	98
68) m/p-Xylenes	10.305	106	234909	38.348	ug/l	96
69) o-Xylene	10.640	106	117314	19.208	ug/l	97
70) Styrene	10.659	104	196552	19.865	ug/l	99
71) Bromoform	10.799	173	51918	19.199	ug/l	98
73) Isopropylbenzene	10.963	105	304993	19.810	ug/l	100
74) N-amyl acetate	10.841	43	107333	20.170	ug/l	95
75) 1,1,2,2-Tetrachloroethane	11.213	83	111776	22.373	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	111878	25.378	ug/l	90
77) Bromobenzene	11.201	156	79635	19.861	ug/l	92
78) n-propylbenzene	11.305	91	343661	20.130	ug/l	99
79) 2-Chlorotoluene	11.366	91	207301	19.531	ug/l	98
80) 1,3,5-Trimethylbenzene	11.451	105	257168	19.678	ug/l	97
81) trans-1,4-Dichloro-2-b...	11.018	75	25223	17.003	ug/l #	84
82) 4-Chlorotoluene	11.457	91	240017	19.631	ug/l	95
83) tert-Butylbenzene	11.719	119	263103	19.715	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	256542	19.747	ug/l	98
85) sec-Butylbenzene	11.890	105	313526	19.473	ug/l	99
86) p-Isopropyltoluene	12.012	119	265785	19.135	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	151410	20.096	ug/l	98
88) 1,4-Dichlorobenzene	12.042	146	153901	19.455	ug/l	99
89) n-Butylbenzene	12.335	91	215724	18.915	ug/l	99
90) Hexachloroethane	12.542	117	42988	18.390	ug/l	97
91) 1,2-Dichlorobenzene	12.335	146	157938	21.102	ug/l	97
92) 1,2-Dibromo-3-Chloropr...	12.945	75	22662	20.396	ug/l	96
93) 1,2,4-Trichlorobenzene	13.591	180	94369	21.199	ug/l	99
94) Hexachlorobutadiene	13.725	225	40165	20.220	ug/l	93
95) Naphthalene	13.780	128	326220	21.998	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	95002	21.742	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050622\
Data File : VX028542.D
Acq On : 06 May 2022 12:16
Operator : JC/MD
Sample : VX0506WBSD01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :
VX0506WBSD01


Quant Time: May 07 05:59:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: May 07 05:59:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16




Instrument :
MSVOA_X
ClientSampleId :
VX0506WBSD01

Tgt Ion:168 Resp: 288852
Ion Ratio Lower Upper
168 100
99 53.1 41.7 62.5



#2
Dichlorodifluoromethane
Concen: 22.678 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16

Tgt Ion: 85 Resp: 42286
Ion Ratio Lower Upper
85 100
87 31.7 16.2 48.6









#3
Chloromethane
Concen: 21.000 ug/l
RT: 1.294 min Scan# 34
Delta R.T. 0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16




Instrument :
MSVOA_X
ClientSampleId :
VX0506WBSD01

Tgt Ion: 50 Resp: 42106
Ion Ratio Lower Upper
50 100
52 32.1 26.4 39.6



#4
Vinyl Chloride
Concen: 17.346 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16

Tgt Ion: 62 Resp: 50512
Ion Ratio Lower Upper
62 100
64 31.8 25.5 38.3





#5

Bromomethane

Concen: 18.130 ug/l

RT: 1.605 min Scan# 84

Delta R.T. 0.006 min

Lab File: VX028542.D

Acq: 06 May 2022 12:16

Instrument :

MSVOA_X

ClientSampleId :





VX0506WBSD01

Tgt Ion: 94 Resp: 60573

Ion Ratio Lower Upper

94 100

96 97.1 75.3 112.9



#6

Chloroethane

Concen: 15.953 ug/l

RT: 1.685 min Scan# 98

Delta R.T. 0.007 min

Lab File: VX028542.D




Acq: 06 May 2022 12:16


Tgt Ion: 64 Resp: 37845

Ion Ratio Lower Upper

64 100

66 30.6 25.3 37.9









#7
 Trichlorofluoromethane
 Concen: 14.990 ug/l
 RT: 1.886 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX028542.D
 Acq: 06 May 2022 12:16




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0506WBSD01


Tgt Ion:101 Resp: 96257
 Ion Ratio Lower Upper
 101 100
 103 65.5 50.5 75.7



#8
 Diethyl Ether
 Concen: 17.677 ug/l
 RT: 2.136 min Scan# 172
 Delta R.T. 0.006 min
 Lab File: VX028542.D
 Acq: 06 May 2022 12:16

Tgt Ion: 74 Resp: 37992
 Ion Ratio Lower Upper
 74 100
 45 80.8 42.0 126.0









#9
1,1,2-Trichlorotrifluoroethane
Concen: 22.500 ug/l
RT: 2.331 min Scan# 204
Delta R.T. -0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16




Instrument :
MSVOA_X
ClientSampleId :
VX0506WBSD01


Tgt Ion:101 Resp: 52630
Ion Ratio Lower Upper
101 100
85 44.0 33.5 50.3
151 77.9 61.8 92.6



#10
Methyl Iodide
Concen: 20.992 ug/l
RT: 2.453 min Scan# 224
Delta R.T. -0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16

Tgt Ion:142 Resp: 56300
Ion Ratio Lower Upper
142 100
127 43.1 33.9 50.9
141 14.3 11.4 17.2









#11
Tert butyl alcohol
Concen: 87.655 ug/l
RT: 2.959 min Scan# 308
Delta R.T. -0.006 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16




Instrument :
MSVOA_X
ClientSampleId :
VX0506WBSD01


Tgt Ion: 59 Resp: 67876
Ion Ratio Lower Upper
59 100
57 9.8 8.3 12.5



#12
1,1-Dichloroethene
Concen: 20.382 ug/l
RT: 2.319 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16

Tgt Ion: 96 Resp: 45246
Ion Ratio Lower Upper
96 100
61 148.2 124.6 187.0
98 64.1 51.0 76.4









#13
Acrolein
Concen: 125.114 ug/l
RT: 2.233 min Scan# 188
Delta R.T. 0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16




Instrument :
MSVOA_X
ClientSampleId :
VX0506WBSD01

Tgt Ion: 56 Resp: 57750
Ion Ratio Lower Upper
56 100
55 70.1 55.0 82.6



#14
Allyl chloride
Concen: 17.842 ug/l
RT: 2.666 min Scan# 259
Delta R.T. 0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16

Tgt Ion: 41 Resp: 66460
Ion Ratio Lower Upper
41 100
39 72.3 50.4 75.6
76 39.9 32.7 49.1









#15
Acrylonitrile
Concen: 104.637 ug/l
RT: 3.062 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16




Instrument :
MSVOA_X
ClientSampleId :
VX0506WBSD01


Tgt Ion: 53 Resp: 160601
Ion Ratio Lower Upper
53 100
52 81.4 65.7 98.5
51 35.7 28.4 42.6



#16
Acetone
Concen: 111.963 ug/l
RT: 2.380 min Scan# 212
Delta R.T. -0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16

Tgt Ion: 43 Resp: 129683
Ion Ratio Lower Upper
43 100
58 31.6 28.8 43.2









#17
Carbon Disulfide
Concen: 15.324 ug/l
RT: 2.514 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16




Instrument :
MSVOA_X
ClientSampleId :
VX0506WBSD01


Tgt Ion: 76 Resp: 91610
Ion Ratio Lower Upper
76 100
78 9.4 7.2 10.8



#18
Methyl Acetate
Concen: 20.782 ug/l
RT: 2.703 min Scan# 265
Delta R.T. -0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16

Tgt Ion: 43 Resp: 69593
Ion Ratio Lower Upper
43 100
74 28.2 22.8 34.2









#19
Methyl tert-butyl Ether
Concen: 20.481 ug/l
RT: 3.111 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16




Instrument :
MSVOA_X
ClientSampleId :
VX0506WBSD01

Tgt Ion: 73 Resp: 189588
Ion Ratio Lower Upper
73 100
57 20.0 16.8 25.2



#20
Methylene Chloride
Concen: 20.956 ug/l
RT: 2.788 min Scan# 279
Delta R.T. -0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16

Tgt Ion: 84 Resp: 58562
Ion Ratio Lower Upper
84 100
49 109.1 85.5 128.3
51 35.2 26.6 39.8
86 65.4 51.5 77.3









#21
trans-1,2-Dichloroethene
Concen: 17.418 ug/l
RT: 3.093 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16




Instrument :
MSVOA_X
ClientSampleId :
VX0506WBSD01


Tgt Ion: 96 Resp: 52584
Ion Ratio Lower Upper
96 100
61 118.7 105.4 158.2
98 59.9 50.0 75.0



#22
Diisopropyl ether
Concen: 19.427 ug/l
RT: 3.757 min Scan# 438
Delta R.T. -0.006 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16

Tgt Ion: 45 Resp: 164816
Ion Ratio Lower Upper
45 100
43 88.0 50.8 76.2#
87 32.8 26.6 39.8
59 12.3 10.5 15.7









#23
 Vinyl Acetate
 Concen: 95.149 ug/l
 RT: 3.721 min Scan# 41
 Delta R.T. -0.000 min
 Lab File: VX028542.D
 Acq: 06 May 2022 12:16




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0506WBSD01


Tgt Ion: 43 Resp: 705248
 Ion Ratio Lower Upper
 43 100
 86 13.6 10.8 16.2



#24
 1,1-Dichloroethane
 Concen: 19.380 ug/l
 RT: 3.611 min Scan# 414
 Delta R.T. -0.000 min
 Lab File: VX028542.D
 Acq: 06 May 2022 12:16

Tgt Ion: 63 Resp: 97086
 Ion Ratio Lower Upper
 63 100
 98 8.4 3.9 11.7
 100 4.8 2.4 7.2









#25
2-Butanone
Concen: 94.075 ug/l
RT: 4.556 min Scan# 569
Delta R.T. -0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16




Instrument :
MSVOA_X
ClientSampleId :
VX0506WBSD01


Tgt Ion: 43 Resp: 209839
Ion Ratio Lower Upper
43 100
72 30.1 24.4 36.6



#26
2,2-Dichloropropane
Concen: 15.910 ug/l
RT: 4.477 min Scan# 556
Delta R.T. -0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16

Tgt Ion: 77 Resp: 77816
Ion Ratio Lower Upper
77 100
97 23.9 11.8 35.4





#27

cis-1,2-Dichloroethene

Concen: 18.730 ug/l

RT: 4.489 min Scan# 51

Delta R.T. -0.000 min

Lab File: VX028542.D

Acq: 06 May 2022 12:16

Instrument :

MSVOA_X

ClientSampleId :

VX0506WBSD01





Tgt Ion: 96 Resp: 67425

Ion Ratio Lower Upper

96 100

61 128.4 0.0 283.8

98 66.3 0.0 129.8



#28

Bromochloromethane

Concen: 22.571 ug/l

RT: 4.903 min Scan# 626

Delta R.T. 0.006 min

Lab File: VX028542.D

Acq: 06 May 2022 12:16




Tgt Ion: 49 Resp: 48510


Ion Ratio Lower Upper

49 100

129 2.3 0.0 4.4

130 96.9 70.7 106.1





#29

Tetrahydrofuran

Concen: 89.824 ug/l

RT: 5.007 min Scan# 64

Delta R.T. 0.000 min

Lab File: VX028542.D

Acq: 06 May 2022 12:16

Instrument :

MSVOA_X

ClientSampleId :

VX0506WBSD01





Tgt Ion: 42 Resp: 130639

Ion Ratio Lower Upper

42 100

72 55.6 44.1 66.1

71 51.9 40.5 60.7



#30

Chloroform

Concen: 18.912 ug/l

RT: 5.092 min Scan# 657

Delta R.T. -0.007 min

Lab File: VX028542.D




Acq: 06 May 2022 12:16

Tgt Ion: 83 Resp: 113778

Ion Ratio Lower Upper

83 100


85 60.6 52.7 79.1








#31
Cyclohexane
Concen: 15.509 ug/l
RT: 5.470 min Scan# 719
Delta R.T. 0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16

Instrument :
MSVOA_X
ClientSampleId :
VX0506WBSD01







Tgt Ion: 56 Resp: 71166
Ion Ratio Lower Upper
56 100
69 33.4 26.7 40.1
84 100.8 77.2 115.8



#32
1,1,1-Trichloroethane
Concen: 17.519 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16

Tgt Ion: 97 Resp: 95438
Ion Ratio Lower Upper
97 100
99 64.1 51.8 77.6
61 40.9 34.0 51.0





#33

1,2-Dichloroethane-d4

Concen: 52.347 ug/l

RT: 5.958 min Scan# 799

Delta R.T. 0.000 min

Lab File: VX028542.D

Acq: 06 May 2022 12:16

Instrument :

MSVOA_X

ClientSampleId :





VX0506WBSD01

Tgt Ion: 65 Resp: 197154

Ion Ratio Lower Upper

65 100

67 51.7 0.0 107.4



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. -0.000 min

Lab File: VX028542.D

Acq: 06 May 2022 12:16




Tgt Ion: 114 Resp: 451402


Ion Ratio Lower Upper

114 100

63 18.2 0.0 36.8

88 15.2 0.0 32.4





#35

Dibromofluoromethane

Concen: 55.698 ug/l

RT: 5.385 min Scan# 705

Delta R.T. 0.000 min

Lab File: VX028542.D

Acq: 06 May 2022 12:16

Instrument :

MSVOA_X

ClientSampleId :

VX0506WBSD01





Tgt Ion:113 Resp: 176320

Ion Ratio Lower Upper

113 100

111 103.6 82.8 124.2

192 20.9 16.9 25.3



#36

1,1-Dichloropropene

Concen: 17.789 ug/l

RT: 5.696 min Scan# 756

Delta R.T. -0.000 min

Lab File: VX028542.D

Acq: 06 May 2022 12:16




Tgt Ion: 75 Resp: 73455


Ion Ratio Lower Upper

75 100

110 37.2 18.3 54.8

77 30.6 24.8 37.2





#37

Ethyl Acetate

Concen: 19.139 ug/l

RT: 4.714 min Scan# 594

Delta R.T. 0.006 min

Lab File: VX028542.D

Acq: 06 May 2022 12:16

Instrument :

MSVOA_X

ClientSampleId :

VX0506WBSD01





Tgt Ion: 43 Resp: 80193

Ion Ratio Lower Upper

43 100

61 15.9 12.6 19.0

70 12.9 10.2 15.2



#38

Carbon Tetrachloride

Concen: 17.764 ug/l

RT: 5.678 min Scan# 753

Delta R.T. -0.006 min

Lab File: VX028542.D

Acq: 06 May 2022 12:16




Tgt Ion: 117 Resp: 82807


Ion Ratio Lower Upper

117 100

119 96.5 77.3 115.9

121 31.1 24.2 36.2





#39

Methylcyclohexane

Concen: 15.687 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. -0.000 min

Lab File: VX028542.D

Acq: 06 May 2022 12:16

Instrument :

MSVOA_X

ClientSampleId :

VX0506WBSD01





Tgt Ion: 83 Resp: 81734

Ion Ratio Lower Upper

83 100

55 68.7 54.6 81.8

98 52.1 37.4 56.2



#40

Benzene

Concen: 18.767 ug/l

RT: 6.043 min Scan# 813

Delta R.T. -0.001 min

Lab File: VX028542.D




Acq: 06 May 2022 12:16

Tgt Ion: 78 Resp: 226842

Ion Ratio Lower Upper

78 100

77 22.9 19.3 28.9








#41
Methacrylonitrile
Concen: 19.819 ug/l
RT: 4.916 min Scan# 61
Delta R.T. -0.006 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16




Instrument :
MSVOA_X
ClientSampleId :
VX0506WBSD01


Tgt Ion:	41	Resp:	42844
Ion	Ratio	Lower	Upper
41	100		
39	54.5	43.0	64.4
67	86.8	70.2	105.4
52	34.2	26.0	39.0



#42
1,2-Dichloroethane
Concen: 18.865 ug/l
RT: 6.092 min Scan# 821
Delta R.T. 0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16


Tgt Ion:	62	Resp:	84644
Ion	Ratio	Lower	Upper
62	100		
98	10.3	0.0	20.4








#43
Isopropyl Acetate
Concen: 18.727 ug/l
RT: 6.342 min Scan# 861
Delta R.T. 0.006 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16

Instrument :
MSVOA_X
ClientSampleId :
VX0506WBSD01







Tgt Ion: 43 Resp: 127704
Ion Ratio Lower Upper
43 100
61 21.4 18.5 27.7
87 16.1 13.0 19.6



#44
Trichloroethene
Concen: 18.938 ug/l
RT: 7.129 min Scan# 991
Delta R.T. -0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16

Tgt Ion:130 Resp: 66381
Ion Ratio Lower Upper
130 100
95 91.0 0.0 188.8





#45

1,2-Dichloropropane

Concen: 19.924 ug/l

RT: 7.433 min Scan# 1041

Delta R.T. -0.001 min

Lab File: VX028542.D

Acq: 06 May 2022 12:16

Instrument :

MSVOA_X

ClientSampleId :





VX0506WBSD01

Tgt Ion: 63 Resp: 58949

Ion Ratio Lower Upper

63 100

65 30.3 25.8 38.6



#46

Dibromomethane

Concen: 19.818 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. -0.000 min

Lab File: VX028542.D

Acq: 06 May 2022 12:16




Tgt Ion: 93 Resp: 45996

Ion Ratio Lower Upper

93 100

95 83.2 65.8 98.6

174 102.9 85.4 128.2









#47
Bromodichloromethane
Concen: 19.282 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16




Instrument :
MSVOA_X
ClientSampleId :
VX0506WBSD01


Tgt Ion: 83 Resp: 86012
Ion Ratio Lower Upper
83 100
85 62.1 52.2 78.2
127 8.8 6.8 10.2



#48
Methyl methacrylate
Concen: 19.149 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16

Tgt Ion: 41 Resp: 59951
Ion Ratio Lower Upper
41 100
69 99.0 82.5 123.7
39 63.6 43.8 65.8





#49

1,4-Dioxane

Concen: 370.169 ug/l

RT: 7.665 min Scan# 1078

Delta R.T. 0.006 min

Lab File: VX028542.D

Acq: 06 May 2022 12:16

Instrument :

MSVOA_X

ClientSampleId :

VX0506WBSD01





Tgt Ion: 88 Resp: 31003

Ion Ratio Lower Upper

88 100

43 29.4 21.9 32.9

58 61.2 53.7 80.5



#50

Toluene-d8

Concen: 50.221 ug/l

RT: 8.653 min Scan# 1241

Delta R.T. -0.000 min

Lab File: VX028542.D




Acq: 06 May 2022 12:16


Tgt Ion: 98 Resp: 610614

Ion Ratio Lower Upper

98 100

100 66.1 52.6 78.8





#51

4-Methyl-2-Pentanone

Concen: 97.686 ug/l

RT: 8.574 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX028542.D

Acq: 06 May 2022 12:16

Instrument :

MSVOA_X

ClientSampleId :





VX0506WBSD01

Tgt Ion: 43 Resp: 427467

Ion Ratio Lower Upper

43 100

58 43.4 36.1 54.1



#52

Toluene

Concen: 18.086 ug/l

RT: 8.720 min Scan# 1252

Delta R.T. -0.000 min

Lab File: VX028542.D




Acq: 06 May 2022 12:16


Tgt Ion: 92 Resp: 151878

Ion Ratio Lower Upper

92 100

91 170.0 137.5 206.3





#53

t-1,3-Dichloropropene

Concen: 17.455 ug/l

RT: 8.982 min Scan# 1295

Delta R.T. -0.000 min

Lab File: VX028542.D

Acq: 06 May 2022 12:16

Instrument :

MSVOA_X

ClientSampleId :





VX0506WBSD01

Tgt Ion: 75 Resp: 82748

Ion Ratio Lower Upper

75 100

77 31.0 25.4 38.0



#54

cis-1,3-Dichloropropene

Concen: 18.155 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX028542.D

Acq: 06 May 2022 12:16




Tgt Ion: 75 Resp: 92422

Ion Ratio Lower Upper

75 100

77 32.3 24.8 37.2

39 42.4 29.6 44.4









#55
1,1,2-Trichloroethane
Concen: 20.246 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. -0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16




Instrument :
MSVOA_X
ClientSampleId :
VX0506WBSD01

Tgt Ion: 97	Resp: 69077
Ion Ratio	Lower Upper
97 100	
83 83.9	66.2 99.4
85 52.7	42.9 64.3
99 66.0	49.0 73.6



#56
Ethyl methacrylate
Concen: 19.296 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16

Tgt Ion: 69	Resp: 95435
Ion Ratio	Lower Upper
69 100	
41 57.0	43.8 65.6
39 36.3	22.2 33.2#





#57

1,3-Dichloropropane

Concen: 19.878 ug/l

RT: 9.311 min Scan# 11349

Delta R.T. 0.000 min

Lab File: VX028542.D

Acq: 06 May 2022 12:16

Instrument :

MSVOA_X

ClientSampleId :





VX0506WBSD01

Tgt Ion: 76 Resp: 109467

Ion Ratio Lower Upper

76 100

78 31.9 25.8 38.8



#58

2-Chloroethyl Vinyl ether

Concen: 98.953 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. 0.000 min

Lab File: VX028542.D




Acq: 06 May 2022 12:16


Tgt Ion: 63 Resp: 229627

Ion Ratio Lower Upper

63 100

106 32.6 23.9 35.9









#59
2-Hexanone
Concen: 96.927 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. 0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16




Instrument :
MSVOA_X
ClientSampleId :
VX0506WBSD01


Tgt Ion: 43 Resp: 328550
Ion Ratio Lower Upper
43 100
58 59.7 31.9 95.7



#60
Dibromochloromethane
Concen: 19.313 ug/l
RT: 9.525 min Scan# 1384
Delta R.T. -0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16

Tgt Ion: 129 Resp: 71837
Ion Ratio Lower Upper
129 100
127 77.6 40.3 120.8









#61
1,2-Dibromoethane
Concen: 19.385 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. -0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16




Instrument :
MSVOA_X
ClientSampleId :
VX0506WBSD01


Tgt Ion:107 Resp: 72561
Ion Ratio Lower Upper
107 100
109 92.0 74.4 111.6



#62
4-Bromofluorobenzene
Concen: 50.701 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16

Tgt Ion: 95 Resp: 235704
Ion Ratio Lower Upper
95 100
174 86.3 0.0 163.0
176 84.3 0.0 155.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16




Instrument :
MSVOA_X
ClientSampleId :
VX0506WBSD01


Tgt Ion:117 Resp: 420997
Ion Ratio Lower Upper
117 100
82 52.0 41.8 62.8
119 32.6 24.8 37.2



#64
Tetrachloroethene
Concen: 19.238 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16

Tgt Ion:164 Resp: 63261
Ion Ratio Lower Upper
164 100
166 129.2 101.2 151.8
129 92.2 73.0 109.4
131 86.0 74.2 111.2









#65
Chlorobenzene
Concen: 19.564 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16




Instrument :
MSVOA_X
ClientSampleId :
VX0506WBSD01


Tgt Ion:112 Resp: 175608
Ion Ratio Lower Upper
112 100
114 31.2 25.8 38.8



#66
1,1,1,2-Tetrachloroethane
Concen: 20.086 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16

Tgt Ion:131 Resp: 67163
Ion Ratio Lower Upper
131 100
133 95.0 48.3 144.8
119 64.4 32.6 98.0









#67
Ethyl Benzene
Concen: 19.479 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16




Instrument :
MSVOA_X
ClientSampleId :
VX0506WBSD01


Tgt Ion: 91 Resp: 293181
Ion Ratio Lower Upper
91 100
106 31.7 24.6 36.8



#68
m/p-Xylenes
Concen: 38.348 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. -0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16

Tgt Ion: 106 Resp: 234909
Ion Ratio Lower Upper
106 100
91 194.5 160.4 240.6









#69
o-Xylene
Concen: 19.208 ug/l
RT: 10.640 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16




Instrument :
MSVOA_X
ClientSampleId :
VX0506WBSD01


Tgt Ion:106 Resp: 117314
Ion Ratio Lower Upper
106 100
91 206.6 106.1 318.1



#70
Styrene
Concen: 19.865 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. -0.001 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16

Tgt Ion:104 Resp: 196552
Ion Ratio Lower Upper
104 100
78 47.2 38.6 57.8
103 54.7 43.8 65.8









#71
Bromoform
Concen: 19.199 ug/l
RT: 10.799 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16




Instrument :
MSVOA_X
ClientSampleId :
VX0506WBSD01


Tgt Ion:173 Resp: 51918
Ion Ratio Lower Upper
173 100
175 49.4 24.1 72.2
254 0.1 0.1 0.1



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16

Tgt Ion:152 Resp: 223996
Ion Ratio Lower Upper
152 100
115 66.3 37.5 112.7
150 163.5 0.0 338.2





#73

Isopropylbenzene

Concen: 19.810 ug/l

RT: 10.963 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX028542.D

Acq: 06 May 2022 12:16

Instrument :

MSVOA_X

ClientSampleId :





VX0506WBSD01

Tgt Ion:105 Resp: 304993

Ion Ratio Lower Upper

105 100

120 27.1 13.5 40.5



#74

N-ethyl acetate

Concen: 20.170 ug/l

RT: 10.841 min Scan# 1600

Delta R.T. -0.007 min

Lab File: VX028542.D

Acq: 06 May 2022 12:16

Tgt Ion: 43 Resp: 107333




Ion Ratio Lower Upper


43 100

70 48.9 42.5 63.7

55 26.6 23.1 34.7

61 26.2 22.8 34.2









#75
1,1,2,2-Tetrachloroethane
Concen: 22.373 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. 0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16




Instrument :
MSVOA_X
ClientSampleId :
VX0506WBSD01


Tgt Ion: 83 Resp: 111776
Ion Ratio Lower Upper
83 100
131 11.0 5.1 15.3
85 64.7 32.6 97.7



#76
1,2,3-Trichloropropane
Concen: 25.378 ug/l
RT: 11.238 min Scan# 1665
Delta R.T. -0.006 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16

Tgt Ion: 75 Resp: 111878
Ion Ratio Lower Upper
75 100
77 32.5 19.4 58.2









#77
Bromobenzene
Concen: 19.861 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16




Instrument :
MSVOA_X
ClientSampleId :
VX0506WBSD01


Tgt Ion:156 Resp: 79635
Ion Ratio Lower Upper
156 100
77 143.2 78.5 235.3
158 97.5 50.2 150.7



#78
n-propylbenzene
Concen: 20.130 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16

Tgt Ion: 91 Resp: 343661
Ion Ratio Lower Upper
91 100
120 23.7 11.6 34.8









#79
2-Chlorotoluene
Concen: 19.531 ug/l
RT: 11.366 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16




Instrument :
MSVOA_X
ClientSampleId :
VX0506WBSD01


Tgt Ion: 91 Resp: 207301
Ion Ratio Lower Upper
91 100
126 36.0 17.3 51.9



#80
1,3,5-Trimethylbenzene
Concen: 19.678 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16

Tgt Ion: 105 Resp: 257168
Ion Ratio Lower Upper
105 100
120 50.4 24.3 72.9









#81
trans-1,4-Dichloro-2-butene
Concen: 17.003 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16




Instrument :
MSVOA_X
ClientSampleId :
VX0506WBSD01


Tgt Ion: 75 Resp: 25223
Ion Ratio Lower Upper
75 100
53 106.4 69.0 103.4#
89 49.2 37.0 55.6



#82
4-Chlorotoluene
Concen: 19.631 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16

Tgt Ion: 91 Resp: 240017
Ion Ratio Lower Upper
91 100
126 32.2 14.9 44.5





#83

tert-Butylbenzene

Concen: 19.715 ug/l

RT: 11.719 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX028542.D

Acq: 06 May 2022 12:16

Instrument :

MSVOA_X

ClientSampleId :

VX0506WBSD01





Tgt Ion:119 Resp: 263103

Ion Ratio Lower Upper

119 100

91 54.5 26.9 80.6

134 23.8 11.4 34.2



#84

1,2,4-Trimethylbenzene

Concen: 19.747 ug/l

RT: 11.756 min Scan# 1750

Delta R.T. -0.000 min

Lab File: VX028542.D




Acq: 06 May 2022 12:16


Tgt Ion:105 Resp: 256542

Ion Ratio Lower Upper

105 100

120 46.2 22.6 67.8









#85
 sec-Butylbenzene
 Concen: 19.473 ug/l
 RT: 11.890 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX028542.D
 Acq: 06 May 2022 12:16




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0506WBSD01


Tgt Ion:105 Resp: 313526
 Ion Ratio Lower Upper
 105 100
 134 20.7 10.2 30.4



#86
 p-Isopropyltoluene
 Concen: 19.135 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. -0.000 min
 Lab File: VX028542.D
 Acq: 06 May 2022 12:16

Tgt Ion:119 Resp: 265785
 Ion Ratio Lower Upper
 119 100
 134 26.7 13.0 39.0
 91 22.7 11.3 33.9









#87
1,3-Dichlorobenzene
Concen: 20.096 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. 0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16




Instrument :
MSVOA_X
ClientSampleId :
VX0506WBSD01


Tgt Ion:146 Resp: 151410
Ion Ratio Lower Upper
146 100
111 38.6 19.7 59.1
148 64.4 31.6 94.7



#88
1,4-Dichlorobenzene
Concen: 19.455 ug/l
RT: 12.042 min Scan# 1797
Delta R.T. 0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16

Tgt Ion:146 Resp: 153901
Ion Ratio Lower Upper
146 100
111 38.9 19.2 57.6
148 65.5 32.3 96.9





#89

n-Butylbenzene

Concen: 18.915 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.000 min

Lab File: VX028542.D

Acq: 06 May 2022 12:16

Instrument :

MSVOA_X

ClientSampleId :

VX0506WBSD01





Tgt Ion: 91 Resp: 215724

Ion Ratio Lower Upper

91 100

92 54.6 27.8 83.3

134 27.4 13.3 39.8



#90

Hexachloroethane

Concen: 18.390 ug/l

RT: 12.542 min Scan# 1879

Delta R.T. 0.000 min

Lab File: VX028542.D




Acq: 06 May 2022 12:16


Tgt Ion: 117 Resp: 42988

Ion Ratio Lower Upper

117 100

201 76.5 39.5 118.5









#91
1,2-Dichlorobenzene
Concen: 21.102 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16




Instrument :
MSVOA_X
ClientSampleId :
VX0506WBSD01


Tgt Ion:146 Resp: 157938
Ion Ratio Lower Upper
146 100
111 39.5 21.0 63.0
148 63.3 32.6 97.7



#92
1,2-Dibromo-3-Chloropropane
Concen: 20.396 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. -0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16

Tgt Ion: 75 Resp: 22662
Ion Ratio Lower Upper
75 100
155 95.8 47.0 141.2
157 121.1 57.9 173.6





#93

1,2,4-Trichlorobenzene

Concen: 21.199 ug/l

RT: 13.591 min Scan# 2051

Delta R.T. -0.000 min

Lab File: VX028542.D

Acq: 06 May 2022 12:16

Instrument :

MSVOA_X

ClientSampleId :

VX0506WBSD01





Tgt Ion:180 Resp: 94369

Ion Ratio Lower Upper

180 100

182 93.6 47.7 143.1

145 30.4 15.2 45.5



#94

Hexachlorobutadiene

Concen: 20.220 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. 0.000 min

Lab File: VX028542.D

Acq: 06 May 2022 12:16




Tgt Ion:225 Resp: 40165


Ion Ratio Lower Upper

225 100

223 61.0 34.2 102.5

227 64.0 33.9 101.7









#95
Naphthalene
Concen: 21.998 ug/l
RT: 13.780 min Scan# 2082
Delta R.T. -0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16

Instrument :
MSVOA_X
ClientSampleId :
VX0506WBSD01

Tgt Ion:128 Resp: 326220
Ion Ratio Lower Upper
128 100
127 12.6 10.2 15.4
129 10.9 8.7 13.1



#96
1,2,3-Trichlorobenzene
Concen: 21.742 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. -0.000 min
Lab File: VX028542.D
Acq: 06 May 2022 12:16

Tgt Ion:180 Resp: 95002
Ion Ratio Lower Upper
180 100
182 94.6 50.0 149.8
145 31.4 16.2 48.5

