

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050622\
 Data File : VX028540.D
 Acq On : 06 May 2022 11:21
 Operator : JC/MD
 Sample : VX0506WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0506WBL01

Quant Time: May 07 05:58:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	376806	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	640962	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	604352	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	286445	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	257974	52.507	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	105.020%
35) Dibromofluoromethane	5.385	113	245765	54.675	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	109.340%
50) Toluene-d8	8.647	98	870593	50.427	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.860%
62) 4-Bromofluorobenzene	11.079	95	310326	47.010	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050622\
Data File : VX028540.D
Acq On : 06 May 2022 11:21
Operator : JC/MD
Sample : VX0506WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0506WBL01

Quant Time: May 07 05:58:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

