

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050622\
 Data File : VX028545.D
 Acq On : 06 May 2022 13:25
 Operator : JC/MD
 Sample : PB144610ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144610ZHE#02

Quant Time: May 07 06:01:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	391222	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	676657	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	671877	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	313591	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	279663	54.824	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 109.640%			
35) Dibromofluoromethane	5.391	113	260265	54.846	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 109.700%			
50) Toluene-d8	8.653	98	949409	52.091	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 104.180%			
62) 4-Bromofluorobenzene	11.079	95	346700	49.750	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 99.500%			
Target Compounds						
16) Acetone	2.380	43	13006	6.991	ug/l	96
20) Methylene Chloride	2.788	84	10137	2.678	ug/l	96
43) Isopropyl Acetate	6.342	43	28526	2.791	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050622\
Data File : VX028545.D
Acq On : 06 May 2022 13:25
Operator : JC/MD
Sample : PB144610ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144610ZHE#02

Quant Time: May 07 06:01:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

