

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX050622\  
 Data File : VX028549.D  
 Acq On : 06 May 2022 14:59  
 Operator : JC/MD  
 Sample : PB144610ZHE#06  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 12 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 PB144610ZHE#06

Quant Time: May 07 06:02:29 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X041922W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Apr 19 13:38:04 2022  
 Response via : Initial Calibration

| Compound                    | R.T.           | QIon | Response            | Conc   | Units | Dev(Min) |
|-----------------------------|----------------|------|---------------------|--------|-------|----------|
| -----                       |                |      |                     |        |       |          |
| Internal Standards          |                |      |                     |        |       |          |
| 1) Pentafluorobenzene       | 5.556          | 168  | 371472              | 50.000 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.763          | 114  | 652499              | 50.000 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5        | 10.055         | 117  | 643390              | 50.000 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024         | 152  | 294035              | 50.000 | ug/l  | 0.00     |
|                             |                |      |                     |        |       |          |
| System Monitoring Compounds |                |      |                     |        |       |          |
| 33) 1,2-Dichloroethane-d4   | 5.958          | 65   | 264301              | 54.567 | ug/l  | 0.00     |
| Spiked Amount 50.000        | Range 61 - 141 |      | Recovery = 109.140% |        |       |          |
| 35) Dibromofluoromethane    | 5.391          | 113  | 244539              | 53.440 | ug/l  | 0.00     |
| Spiked Amount 50.000        | Range 69 - 133 |      | Recovery = 106.880% |        |       |          |
| 50) Toluene-d8              | 8.653          | 98   | 895109              | 50.930 | ug/l  | 0.00     |
| Spiked Amount 50.000        | Range 65 - 126 |      | Recovery = 101.860% |        |       |          |
| 62) 4-Bromofluorobenzene    | 11.079         | 95   | 324794              | 48.332 | ug/l  | 0.00     |
| Spiked Amount 50.000        | Range 58 - 135 |      | Recovery = 96.660%  |        |       |          |
|                             |                |      |                     |        |       |          |
| Target Compounds            |                |      |                     |        |       |          |
| 16) Acetone                 | 2.380          | 43   | 12839               | 7.324  | ug/l  | 98       |
| 20) Methylene Chloride      | 2.788          | 84   | 23647               | 6.580  | ug/l  | 96       |
| 43) Isopropyl Acetate       | 6.348          | 43   | 27225               | 2.762  | ug/l  | 97       |
| -----                       |                |      |                     |        |       |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX050622\  
Data File : VX028549.D  
Acq On : 06 May 2022 14:59  
Operator : JC/MD  
Sample : PB144610ZHE#06  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 12 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
PB144610ZHE#06

Quant Time: May 07 06:02:29 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X041922W.M  
Quant Title : SW846 8260  
QLast Update : Tue Apr 19 13:38:04 2022  
Response via : Initial Calibration

