

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050622\
 Data File : VX028550.D
 Acq On : 06 May 2022 15:22
 Operator : JC/MD
 Sample : PB144610ZHE#07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144610ZHE#07

Quant Time: May 07 06:02:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	375281	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	652163	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	644090	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	292900	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	270338	55.247	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	110.500%
35) Dibromofluoromethane	5.391	113	252549	55.219	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	110.440%
50) Toluene-d8	8.653	98	915547	52.120	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.240%
62) 4-Bromofluorobenzene	11.079	95	333042	49.585	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.180%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	12516	7.018	ug/l	98
20) Methylene Chloride	2.788	84	24271	6.685	ug/l	97
43) Isopropyl Acetate	6.342	43	27309	2.772	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050622\
Data File : VX028550.D
Acq On : 06 May 2022 15:22
Operator : JC/MD
Sample : PB144610ZHE#07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144610ZHE#07

Quant Time: May 07 06:02:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

