

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050622\
 Data File : VX028552.D
 Acq On : 06 May 2022 16:08
 Operator : JC/MD
 Sample : N2685-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-14B-5-GR

Quant Time: May 07 06:03:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	381129	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	659384	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	640608	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	301343	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	275911	55.521	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 111.040%			
35) Dibromofluoromethane	5.391	113	187696	40.590	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 81.180%			
50) Toluene-d8	8.653	98	924991	52.081	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 104.160%			
62) 4-Bromofluorobenzene	11.079	95	341863	50.341	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 100.680%			
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	39322	24.648	ug/l	93
17) Carbon Disulfide	2.514	76	12828	1.626	ug/l	99
20) Methylene Chloride	2.788	84	15834	4.294	ug/l	98
25) 2-Butanone	4.568	43	7409	2.517	ug/l	99
73) Isopropylbenzene	10.963	105	66229	3.198	ug/l	99
78) n-propylbenzene	11.305	91	27289	1.188	ug/l	96
85) sec-Butylbenzene	11.890	105	33750	1.558	ug/l #	77
95) Naphthalene	13.780	128	21673	1.086	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050622\
Data File : VX028552.D
Acq On : 06 May 2022 16:08
Operator : JC/MD
Sample : N2685-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-14B-5-GR

Quant Time: May 07 06:03:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

