

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050622\
 Data File : VX028553.D
 Acq On : 06 May 2022 16:32
 Operator : JC/MD
 Sample : N2685-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-14B-5

Quant Time: May 07 06:03:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	386837	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	661982	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	645020	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	314210	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	270206	53.571	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 107.140%			
35) Dibromofluoromethane	5.391	113	196163	42.254	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 84.500%			
50) Toluene-d8	8.653	98	903577	50.676	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 101.360%			
62) 4-Bromofluorobenzene	11.079	95	337978	49.574	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 99.140%			
Target Compounds						
16) Acetone	2.380	43	57220	35.947	ug/l	96
17) Carbon Disulfide	2.514	76	13872	1.733	ug/l	96
20) Methylene Chloride	2.788	84	23304	6.227	ug/l	98
25) 2-Butanone	4.574	43	12889	4.315	ug/l	94
73) Isopropylbenzene	10.963	105	51006	2.362	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	27514	1.510	ug/l	95
85) sec-Butylbenzene	11.890	105	27064	1.198	ug/l	80
95) Naphthalene	13.780	128	38298	1.841	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050622\
Data File : VX028553.D
Acq On : 06 May 2022 16:32
Operator : JC/MD
Sample : N2685-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-14B-5

Quant Time: May 07 06:03:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

