

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050622\
 Data File : VX028557.D
 Acq On : 06 May 2022 18:05
 Operator : JC/MD
 Sample : N2723-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MMC-PIT1

Quant Time: May 07 06:05:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	368864	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	644036	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	637539	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	292830	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	272343	56.625	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	113.260%
35) Dibromofluoromethane	5.391	113	249356	55.209	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	110.420%
50) Toluene-d8	8.653	98	917213	52.874	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.740%
62) 4-Bromofluorobenzene	11.085	95	330104	49.768	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.540%
Target Compounds						
						Qvalue
16) Acetone	2.373	43	43419	28.319	ug/l	92
20) Methylene Chloride	2.794	84	10479	2.936	ug/l	90
43) Isopropyl Acetate	6.348	43	28246	2.903	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050622\
Data File : VX028557.D
Acq On : 06 May 2022 18:05
Operator : JC/MD
Sample : N2723-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MMC-PIT1

Quant Time: May 07 06:05:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

