

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050622\
 Data File : VX028559.D
 Acq On : 06 May 2022 18:52
 Operator : JC/MD
 Sample : N2723-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MMC-PIT3

Quant Time: May 07 06:05:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	372907	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	650282	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	655920	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	304605	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	271816	55.903	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	111.800%
35) Dibromofluoromethane	5.391	113	250250	54.875	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	109.740%
50) Toluene-d8	8.653	98	921297	52.599	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.200%
62) 4-Bromofluorobenzene	11.079	95	342169	51.091	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.180%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	42933	27.668	ug/l	99
20) Methylene Chloride	2.788	84	9905	2.746	ug/l	95
43) Isopropyl Acetate	6.342	43	28474	2.899	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050622\
Data File : VX028559.D
Acq On : 06 May 2022 18:52
Operator : JC/MD
Sample : N2723-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MMC-PIT3

Quant Time: May 07 06:05:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

