

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050624\
 Data File : VX041373.D
 Acq On : 06 May 2024 12:48
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: May 07 01:58:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 01 02:38:17 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	188054	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	246345	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	234399	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	130458	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	100309	48.997	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.000%
35) Dibromofluoromethane	5.385	113	80017	47.931	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.860%
50) Toluene-d8	8.646	98	262053	49.595	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.200%
62) 4-Bromofluorobenzene	11.079	95	120142	58.738	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	117.480%
Target Compounds						
16) Acetone	2.391	43	7294	4.246	ug/l	99
44) Trichloroethene	7.128	130	16529	10.172	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050624\
Data File : VX041373.D
Acq On : 06 May 2024 12:48
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: May 07 01:58:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 01 02:38:17 2024
Response via : Initial Calibration

