

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX050624\  
 Data File : VX041374.D  
 Acq On : 06 May 2024 13:11  
 Operator : JC/MD  
 Sample : IBLK  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 IBLK

Quant Time: May 07 01:58:52 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X043024W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed May 01 02:38:17 2024  
 Response via : Initial Calibration

| Compound                    | R.T.           | QIon | Response   | Conc     | Units | Dev(Min)  |
|-----------------------------|----------------|------|------------|----------|-------|-----------|
| -----                       |                |      |            |          |       |           |
| Internal Standards          |                |      |            |          |       |           |
| 1) Pentafluorobenzene       | 5.550          | 168  | 185598     | 50.000   | ug/l  | # 0.00    |
| 34) 1,4-Difluorobenzene     | 6.757          | 114  | 252718     | 50.000   | ug/l  | 0.00      |
| 63) Chlorobenzene-d5        | 10.055         | 117  | 261346     | 50.000   | ug/l  | 0.00      |
| 72) 1,4-Dichlorobenzene-d4  | 12.024         | 152  | 136330     | 50.000   | ug/l  | 0.00      |
|                             |                |      |            |          |       |           |
| System Monitoring Compounds |                |      |            |          |       |           |
| 33) 1,2-Dichloroethane-d4   | 5.952          | 65   | 101002     | 49.989   | ug/l  | 0.00      |
| Spiked Amount 50.000        | Range 78 - 117 |      | Recovery = | 99.980%  |       |           |
| 35) Dibromofluoromethane    | 5.385          | 113  | 83134      | 48.543   | ug/l  | 0.00      |
| Spiked Amount 50.000        | Range 75 - 124 |      | Recovery = | 97.080%  |       |           |
| 50) Toluene-d8              | 8.647          | 98   | 288164     | 53.162   | ug/l  | 0.00      |
| Spiked Amount 50.000        | Range 92 - 112 |      | Recovery = | 106.320% |       |           |
| 62) 4-Bromofluorobenzene    | 11.079         | 95   | 121905     | 58.097   | ug/l  | 0.00      |
| Spiked Amount 50.000        | Range 83 - 123 |      | Recovery = | 116.200% |       |           |
|                             |                |      |            |          |       |           |
| Target Compounds            |                |      |            |          |       |           |
| 44) Trichloroethene         | 7.129          | 130  | 7540       | 4.523    | ug/l  | Qvalue 82 |
| -----                       |                |      |            |          |       |           |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX050624\  
Data File : VX041374.D  
Acq On : 06 May 2024 13:11  
Operator : JC/MD  
Sample : IBLK  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
IBLK

Quant Time: May 07 01:58:52 2024  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X043024W.M  
Quant Title : SW846 8260  
QLast Update : Wed May 01 02:38:17 2024  
Response via : Initial Calibration

