

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050624\
 Data File : VX041385.D
 Acq On : 06 May 2024 17:35
 Operator : JC/MD
 Sample : P2393-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-6

Quant Time: May 07 02:02:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 01 02:38:17 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	165686	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	243017	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	230425	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	119496	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	86783	48.113	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.220%
35) Dibromofluoromethane	5.385	113	71004	43.115	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	86.220%
50) Toluene-d8	8.647	98	264866	50.814	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.620%
62) 4-Bromofluorobenzene	11.079	95	99424	49.274	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.540%
Target Compounds						
19) Methyl tert-butyl Ether	3.123	73	4787	1.023	ug/l	92
44) Trichloroethene	7.141	130	528	0.329	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050624\
Data File : VX041385.D
Acq On : 06 May 2024 17:35
Operator : JC/MD
Sample : P2393-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-6

Quant Time: May 07 02:02:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 01 02:38:17 2024
Response via : Initial Calibration

