

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050624\
 Data File : VX041389.D
 Acq On : 06 May 2024 19:07
 Operator : JC/MD
 Sample : PB160704TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB160704TB

Quant Time: May 07 02:05:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 01 02:38:17 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	181141	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	277398	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	255210	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	114936	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	95553	48.456	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	96.920%		
35) Dibromofluoromethane	5.385	113	78681	41.855	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	83.720%		
50) Toluene-d8	8.647	98	297009	49.919	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	99.840%		
62) 4-Bromofluorobenzene	11.079	95	95986	41.674	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	83.340%		
Target Compounds						
						Qvalue
16) Acetone	2.392	43	28194	29.010	ug/l	95
20) Methylene Chloride	2.788	84	5787	3.380	ug/l #	92
25) 2-Butanone	4.586	43	10525	8.246	ug/l	92
37) Ethyl Acetate	4.745	43	3069	1.094	ug/l #	86
43) Isopropyl Acetate	6.342	43	116966	27.737	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050624\
Data File : VX041389.D
Acq On : 06 May 2024 19:07
Operator : JC/MD
Sample : PB160704TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB160704TB

Quant Time: May 07 02:05:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 01 02:38:17 2024
Response via : Initial Calibration

