

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050624\
 Data File : VX041390.D
 Acq On : 06 May 2024 19:30
 Operator : JC/MD
 Sample : P2354-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 V359

Quant Time: May 07 02:05:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 01 02:38:17 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	194996	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	290300	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	272807	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	140582	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	100688	47.432	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.860%
35) Dibromofluoromethane	5.385	113	83878	42.637	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	85.280%
50) Toluene-d8	8.647	98	265068	42.570	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	85.140%#
62) 4-Bromofluorobenzene	11.079	95	120557	50.016	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.040%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	33228	32.137	ug/l	90
18) Methyl Acetate	2.709	43	3907	1.799	ug/l	# 89
20) Methylene Chloride	2.788	84	5465	2.965	ug/l	90
25) 2-Butanone	4.574	43	18542	13.496	ug/l	96
43) Isopropyl Acetate	6.342	43	63344	14.353	ug/l	99
95) Naphthalene	13.774	128	18079	1.877	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050624\
Data File : VX041390.D
Acq On : 06 May 2024 19:30
Operator : JC/MD
Sample : P2354-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
V359

Quant Time: May 07 02:05:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 01 02:38:17 2024
Response via : Initial Calibration

