

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050624\
 Data File : VX041391.D
 Acq On : 06 May 2024 19:54
 Operator : JC/MD
 Sample : P2355-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VB16173

Quant Time: May 07 02:05:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 01 02:38:17 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	182992	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	273018	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	219017	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	112229	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	96874	48.629	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.260%
35) Dibromofluoromethane	5.385	113	78701	42.537	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	85.080%
50) Toluene-d8	8.647	98	301381	51.466	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.940%
62) 4-Bromofluorobenzene	11.079	95	95831	42.275	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.540%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	34878	36.416	ug/l	100
18) Methyl Acetate	2.715	43	3813	1.871	ug/l	91
20) Methylene Chloride	2.788	84	6243	3.610	ug/l	87
25) 2-Butanone	4.574	43	23722	18.398	ug/l	99
43) Isopropyl Acetate	6.348	43	90308	21.759	ug/l	97
95) Naphthalene	13.780	128	7733	1.006	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050624\
Data File : VX041391.D
Acq On : 06 May 2024 19:54
Operator : JC/MD
Sample : P2355-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VB16173

Quant Time: May 07 02:05:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 01 02:38:17 2024
Response via : Initial Calibration

