

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050625\
 Data File : VX046055.D
 Acq On : 06 May 2025 11:52
 Operator : JC/MD
 Sample : Q1945-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GB1W

Quant Time: May 07 01:45:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	65315	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	126773	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	116564	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	48139	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63334	52.012	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.020%	
35) Dibromofluoromethane	5.385	113	45739	50.103	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.200%	
50) Toluene-d8	8.647	98	159965	50.627	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.260%	
62) 4-Bromofluorobenzene	11.079	95	59254	48.889	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 97.780%	
Target Compounds						
16) Acetone	2.379	43	1370	2.806	ug/l	Qvalue # 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050625\
Data File : VX046055.D
Acq On : 06 May 2025 11:52
Operator : JC/MD
Sample : Q1945-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GB1W

Quant Time: May 07 01:45:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

