

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050625\
 Data File : VX046059.D
 Acq On : 06 May 2025 13:25
 Operator : JC/MD
 Sample : PB167852TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB167852TB

Quant Time: May 07 01:46:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	69899	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	140840	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	132970	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	56151	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	70146	53.828	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.660%
35) Dibromofluoromethane	5.385	113	50822	50.111	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.220%
50) Toluene-d8	8.647	98	174345	49.667	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.340%
62) 4-Bromofluorobenzene	11.079	95	68151	50.614	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	101.220%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	23354	44.697	ug/l	98
18) Methyl Acetate	2.709	43	1626	1.341	ug/l #	88
20) Methylene Chloride	2.788	84	3448	3.444	ug/l	97
43) Isopropyl Acetate	6.342	43	12650	4.925	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050625\
Data File : VX046059.D
Acq On : 06 May 2025 13:25
Operator : JC/MD
Sample : PB167852TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB167852TB

Quant Time: May 07 01:46:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

