

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050625\
 Data File : VX046060.D
 Acq On : 06 May 2025 13:48
 Operator : JC/MD
 Sample : PB167852ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB167852ZHE#01

Quant Time: May 07 01:46:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	65728	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	129361	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	123343	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	54743	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64875	52.943	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 105.880%		
35) Dibromofluoromethane	5.385	113	45931	49.307	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 98.620%		
50) Toluene-d8	8.647	98	164570	51.043	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 102.080%		
62) 4-Bromofluorobenzene	11.079	95	64123	51.848	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 103.700%		
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	21311	43.375	ug/l	98
18) Methyl Acetate	2.703	43	1636	1.435	ug/l	98
20) Methylene Chloride	2.782	84	3041	3.230	ug/l	96
43) Isopropyl Acetate	6.348	43	12242	5.189	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050625\
Data File : VX046060.D
Acq On : 06 May 2025 13:48
Operator : JC/MD
Sample : PB167852ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB167852ZHE#01

Quant Time: May 07 01:46:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

