

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050625\
 Data File : VX046061.D
 Acq On : 06 May 2025 14:12
 Operator : JC/MD
 Sample : PB167852ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB167852ZHE#02

Quant Time: May 07 01:46:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	64468	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	129620	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	122599	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	52490	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64414	53.594	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.180%	
35) Dibromofluoromethane	5.385	113	46760	50.096	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.200%	
50) Toluene-d8	8.647	98	164273	50.849	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.700%	
62) 4-Bromofluorobenzene	11.079	95	64068	51.700	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 103.400%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	21564	44.748	ug/l	99
18) Methyl Acetate	2.709	43	1759	1.573	ug/l	96
20) Methylene Chloride	2.788	84	3240	3.508	ug/l	87
43) Isopropyl Acetate	6.342	43	11835	5.007	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050625\
Data File : VX046061.D
Acq On : 06 May 2025 14:12
Operator : JC/MD
Sample : PB167852ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB167852ZHE#02

Quant Time: May 07 01:46:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

