

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050625\
 Data File : VX046062.D
 Acq On : 06 May 2025 14:35
 Operator : JC/MD
 Sample : PB167852ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB167852ZHE#03

Quant Time: May 07 01:46:54 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	63845	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	129783	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	122777	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	51096	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64856	54.488	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	108.980%
35) Dibromofluoromethane	5.379	113	46889	50.171	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.340%
50) Toluene-d8	8.647	98	161971	50.073	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.140%
62) 4-Bromofluorobenzene	11.079	95	63000	50.775	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	101.540%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	21534	45.121	ug/l	100
18) Methyl Acetate	2.709	43	2115	1.910	ug/l #	88
20) Methylene Chloride	2.788	84	3188	3.486	ug/l	95
43) Isopropyl Acetate	6.342	43	12049	5.091	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050625\
Data File : VX046062.D
Acq On : 06 May 2025 14:35
Operator : JC/MD
Sample : PB167852ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB167852ZHE#03

Quant Time: May 07 01:46:54 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

