

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050625\
 Data File : VX046063.D
 Acq On : 06 May 2025 14:58
 Operator : JC/MD
 Sample : PB167852ZHE#04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB167852ZHE#04

Quant Time: May 07 01:47:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	64597	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	128748	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	122702	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	51734	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64672	53.701	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 107.400%			
35) Dibromofluoromethane	5.379	113	46222	49.855	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 99.720%			
50) Toluene-d8	8.647	98	162013	50.489	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 100.980%			
62) 4-Bromofluorobenzene	11.079	95	63484	51.576	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 103.160%			
Target Compounds					Qvalue	
16) Acetone	2.380	43	21728	44.998	ug/l	100
18) Methyl Acetate	2.703	43	2049	1.829	ug/l #	88
20) Methylene Chloride	2.782	84	3144	3.398	ug/l	92
43) Isopropyl Acetate	6.342	43	12159	5.179	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050625\
Data File : VX046063.D
Acq On : 06 May 2025 14:58
Operator : JC/MD
Sample : PB167852ZHE#04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB167852ZHE#04

Quant Time: May 07 01:47:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

