

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX050719\
 Data File : VX009378.D
 Acq On : 07 May 2019 15:44
 Operator : JC/SP
 Sample : K2721-01 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS003RW114R-190506

Manual Integrations
 APPROVED

MMDadoda
 5/8/2019 10:10:45 AM

Quant Time: May 08 02:37:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 07 08:24:47 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	185135	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	296319	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	259358	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	111903	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	119350	47.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.44%	
35) Dibromofluoromethane	5.50	113	90710	48.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.40%	
50) Toluene-d8	8.71	98	368535	49.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.00%	
62) 4-Bromofluorobenzene	11.13	95	121927	44.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.24%	
Target Compounds						
65) Chlorobenzene	10.13	112	9164	1.766	ug/l	100
87) 1,3-Dichlorobenzene	12.02	146	1639	0.455	ug/l	96
88) 1,4-Dichlorobenzene	12.10	146	4607m	1.269	ug/l	
91) 1,2-Dichlorobenzene	12.39	146	25570	6.972	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050719\
Data File : VX009378.D
Acq On : 07 May 2019 15:44
Operator : JC/SP
Sample : K2721-01 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS003RW114R-190506

Manual Integrations
APPROVED

MMDadoda
5/8/2019 10:10:45 AM

Quant Time: May 08 02:37:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Tue May 07 08:24:47 2019
Response via : Initial Calibration

