

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX050719\
 Data File : VX009381.D
 Acq On : 07 May 2019 16:54
 Operator : JC/SP
 Sample : K2728-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP BLANK

Quant Time: May 08 03:10:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 07 08:24:47 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	179032	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	284496	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	252383	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	108586	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	116164	47.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.06%	
35) Dibromofluoromethane	5.50	113	87232	48.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.56%	
50) Toluene-d8	8.71	98	357435	49.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.00%	
62) 4-Bromofluorobenzene	11.13	95	121232	46.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.42%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050719\
Data File : VX009381.D
Acq On : 07 May 2019 16:54
Operator : JC/SP
Sample : K2728-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP BLANK

Quant Time: May 08 03:10:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
Quant Title : SW846 8260
QLast Update : Tue May 07 08:24:47 2019
Response via : Initial Calibration

