

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX050719\
 Data File : VX009388.D
 Acq On : 07 May 2019 19:39
 Operator : JC/SP
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC020

Quant Time: May 08 02:26:21 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X050119W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu May 02 06:46:57 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	89	0.00
2 M	Dichlorodifluoromethane	1.594	1.417	11.1	80	0.00
3 M	Chloromethane	2.143	1.821	15.0	77	0.00
4 M	Vinyl Chloride	2.138	1.938	9.4	82	0.00
5 M	Bromomethane	1.355	0.947	30.1#	70	0.00
6 M	Chloroethane	1.402	1.261	10.1	79	0.00
7 M	Trichlorofluoromethane	2.640	2.527	4.3	87	0.00
8 T	Diethyl Ether	1.226	1.195	2.5	89	0.00
9	1,1,2-Trichlorotrifluoroeth	1.634	1.573	3.7	87	0.00
10 M	1,1-Dichloroethene	1.641	1.597	2.7	88	0.00
11	Methyl Iodide	1.784	0.884	50.4#	48	0.00
12	Methyl Acetate	3.091	3.194	-3.3	96	0.00
13 M	Acrolein	0.424	0.322	24.1	74	0.00
14 M	Acrylonitrile	1.341	1.419	-5.8	97	0.00
15 M	Acetone	0.385	0.391	-1.6	90	0.00
16 M	Carbon Disulfide	4.440	4.005	9.8	82	0.00
17	Allyl chloride	3.939	3.700	6.1	86	0.00
18 M	Methylene Chloride	1.983	1.955	1.4	90	0.00
19 M	trans-1,2-Dichloroethene	1.750	1.672	4.5	87	0.00
20 T	Diisopropyl ether	7.758	7.575	2.4	89	0.00
21 M	1,1-Dichloroethane	3.754	3.650	2.8	89	0.00
22 M	cis-1,2-Dichloroethene	2.107	2.012	4.5	87	0.00
23 M	tert-Butyl Alcohol	0.567	0.626	-10.4	105	0.01
24 M	Methyl tert-Butyl Ether	6.422	6.371	0.8	92	0.00
25 M	Chloroform	3.426	3.382	1.3	90	0.00
26	Cyclohexane	3.499	3.209	8.3	83	0.00
27 s	1,2-Dichloroethane-d4	2.489	2.511	-0.9	91	0.00
28 I	1,4-Difluorobenzene	1.000	1.000	0.0	90	0.00
29	1,1-Dichloropropene	0.440	0.407	7.5	84	0.00
30 M	2-Butanone	0.345	0.367	-6.4	97	0.00
31	2,2-Dichloropropane	0.485	0.399	17.7	75	0.00
32 M	1,1,1-Trichloroethane	0.477	0.470	1.5	92	0.00
33 M	Carbon Tetrachloride	0.399	0.385	3.5	89	0.00
34 M	Benzene	1.357	1.317	2.9	88	0.00
35	Methacrylonitrile	0.361	0.376	-4.2	95	0.00
36 M	1,2-Dichloroethane	0.492	0.494	-0.4	91	0.00
37 M	Trichloroethene	0.327	0.311	4.9	88	0.00
38	Methylcyclohexane	0.544	0.484	11.0	82	0.00
39 M	1,2-Dichloropropane	0.386	0.375	2.8	89	0.00
40	Dibromomethane	0.219	0.218	0.5	91	0.00
41 M	Bromodichloromethane	0.448	0.442	1.3	92	0.00
42 M	Vinyl Acetate	1.138	1.137	0.1	90	0.00
43	Ethyl Acetate	0.623	0.666	-6.9	95	0.00
44	Isopropyl Acetate	1.009	1.019	-1.0	94	0.00
45 T	1,4-Dioxane	0.010	0.011#	-10.0	101	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX050719\
 Data File : VX009388.D
 Acq On : 07 May 2019 19:39
 Operator : JC/SP
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC020

Quant Time: May 08 02:26:21 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X050119W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu May 02 06:46:57 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46	Methyl methacrylate	0.504	0.512	-1.6	94	0.00
47	n-amyl Acetate	0.909	0.891	2.0	92	0.00
48 M	t-1,3-Dichloropropene	0.523	0.499	4.6	91	0.00
49 T	cis-1,3-Dichloropropene	0.574	0.546	4.9	89	0.00
50 M	1,1,2-Trichloroethane	0.336	0.345	-2.7	94	0.00
51	Ethyl methacrylate	0.621	0.612	1.4	91	0.00
52	1,3-Dichloropropane	0.596	0.600	-0.7	91	0.00
53 M	Dibromochloromethane	0.335	0.331	1.2	93	0.00
54 M	1,2-Dibromoethane	0.345	0.347	-0.6	93	0.00
55 M	2-Chloroethyl vinyl ether	0.331	0.325	1.8	86	0.00
56 M	Bromoform	0.257	0.250	2.7	95	0.00
57 I	Chlorobenzene-d5	1.000	1.000	0.0	92	0.00
58 M	4-Methyl-2-Pentanone	0.738	0.774	-4.9	97	0.00
59 M	2-Hexanone	0.579	0.606	-4.7	97	0.00
60 S	4-Bromofluorobenzene	0.510	0.505	1.0	90	0.00
61 M	Tetrachloroethene	0.331	0.312	5.7	87	0.00
62 M	Toluene	1.603	1.559	2.7	89	0.00
63 S	Toluene-d8	1.422	1.411	0.8	90	0.00
64 M	Chlorobenzene	0.965	0.931	3.5	90	0.00
65	1,1,1,2-Tetrachloroethane	0.353	0.349	1.1	93	0.00
66 M	Ethyl Benzene	1.786	1.706	4.5	89	0.00
67 M	m/p-Xylenes	0.652	0.614	5.8	88	0.00
68 M	o-Xylene	0.641	0.622	3.0	91	0.00
69 M	Styrene	1.125	1.089	3.2	91	0.00
70	Isopropylbenzene	1.724	1.665	3.4	90	0.00
71 M	1,1,2,2-Tetrachloroethane	0.631	0.647	-2.5	97	0.00
72	1,2,3-Trichloropropane	0.544	0.552	-1.5	94	0.00
73	Bromobenzene	0.426	0.413	3.1	90	0.00
74	n-propylbenzene	1.998	1.874	6.2	88	0.00
75	2-Chlorotoluene	1.198	1.146	4.3	89	0.00
76	1,3,5-Trimethylbenzene	1.461	1.391	4.8	89	0.00
77	t-1,4-Dichloro-2-butene	0.220	0.191	13.2	88	0.00
78	4-Chlorotoluene	1.361	1.275	6.3	89	0.00
79	tert-butylbenzene	1.412	1.334	5.5	88	0.00
80	1,2,4-Trimethylbenzene	1.464	1.372	6.3	88	0.00
81	sec-Butylbenzene	1.678	1.589	5.3	89	0.00
82	p-Isopropyltoluene	1.526	1.420	6.9	88	0.00
83 M	1,3-Dichlorobenzene	0.761	0.718	5.7	91	0.00
84 M	1,4-Dichlorobenzene	0.758	0.712	6.1	91	0.00
85	n-Butylbenzene	1.396	1.233	11.7	86	0.00
86 T	Hexachloroethane	0.260	0.239	8.1	87	0.00
87 M	1,2-Dichlorobenzene	0.766	0.743	3.0	92	0.00
88	1,2-Dibromo-3-Chloropropane	0.147	0.154	-4.8	100	0.00
89	1,2,4-Trichlorobenzene	0.546	0.514	5.9	93	0.00
90	Hexachlorobutadiene	0.273	0.254	7.0	90	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX050719\
Data File : VX009388.D
Acq On : 07 May 2019 19:39
Operator : JC/SP
Sample : VSTDCCC020
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC020

Quant Time: May 08 02:26:21 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X050119W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu May 02 06:46:57 2019
Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
91 M	Naphthalene	1.787	1.818	-1.7	96	0.00
92	1,2,3-Trichlorobenzene	0.555	0.539	2.9	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0