


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX050719\
 Data File : VX009366.D
 Acq On : 07 May 2019 09:09
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042919W.M
 Title : SW846 8260
 Last Update : Wed May 08 03:31:13 2019



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1629

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.6	6478	PASS
75	95	30	60	48.4	15227	PASS
95	95	100	100	100.0	31450	PASS
96	95	5	9	6.3	1994	PASS
173	174	0.00	2	0.3	68	PASS
174	95	50	100	79.9	25141	PASS
175	174	5	9	7.6	1900	PASS
176	174	95	101	98.4	24746	PASS
177	176	5	9	6.3	1557	PASS