

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX050720\
 Data File : VX016087.D
 Acq On : 07 May 2020 12:39
 Operator : JC/SP
 Sample : L2543-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z3-TB13-200503

Quant Time: May 08 03:48:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	317502	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	6.83	114	520303	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	498062	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	254932	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	191149	45.75	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.50%	
35) Dibromofluoromethane	5.46	113	151429	45.94	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.88%	
50) Toluene-d8	8.70	98	636838	50.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.28%	
62) 4-Bromofluorobenzene	11.12	95	249284	51.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.20%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050720\
Data File : VX016087.D
Acq On : 07 May 2020 12:39
Operator : JC/SP
Sample : L2543-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z3-TB13-200503

Quant Time: May 08 03:48:06 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

