

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX050720\
 Data File : VX016089.D
 Acq On : 07 May 2020 13:25
 Operator : JC/SP
 Sample : L2543-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS038RW016-200503

Quant Time: May 08 04:57:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	296320	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	479601	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	462357	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	231771	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	178851	45.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.74%	
35) Dibromofluoromethane	5.47	113	140043	46.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.18%	
50) Toluene-d8	8.70	98	583067	49.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.60%	
62) 4-Bromofluorobenzene	11.12	95	233450	52.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.86%	
Target Compounds						
4) Vinyl Chloride	1.39	62	15037	3.869	ug/l	99
16) Acetone	2.42	43	2262	1.300	ug/l	98
27) cis-1,2-Dichloroethene	4.57	96	21270	5.668	ug/l	83
65) Chlorobenzene	10.12	112	119793	12.374	ug/l	97
83) tert-Butylbenzene	11.76	119	19145	1.561	ug/l	98
85) sec-Butylbenzene	11.93	105	34480	2.082	ug/l #	56
91) 1,2-Dichlorobenzene	12.38	146	2341	0.311	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050720\
Data File : VX016089.D
Acq On : 07 May 2020 13:25
Operator : JC/SP
Sample : L2543-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS038RW016-200503

Quant Time: May 08 04:57:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

