

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX050720\
 Data File : VX016094.D
 Acq On : 07 May 2020 15:18
 Operator : JC/SP
 Sample : L2551-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-200430

Quant Time: May 08 04:01:47 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	286725	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	461875	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	454970	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	243177	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	172189	45.64	ug/l	0.00
Spiked Amount			Recovery	=	91.28%	
35) Dibromofluoromethane	5.47	113	133624	45.67	ug/l	0.00
Spiked Amount			Recovery	=	91.34%	
50) Toluene-d8	8.70	98	569296	50.49	ug/l	0.00
Spiked Amount			Recovery	=	100.98%	
62) 4-Bromofluorobenzene	11.12	95	232047	54.11	ug/l	0.00
Spiked Amount			Recovery	=	108.22%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050720\
Data File : VX016094.D
Acq On : 07 May 2020 15:18
Operator : JC/SP
Sample : L2551-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
TB-01-200430

Quant Time: May 08 04:01:47 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

