

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX050720\
 Data File : VX016095.D
 Acq On : 07 May 2020 15:41
 Operator : JC/SP
 Sample : L2550-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z5-S1EFF-200430

Quant Time: May 08 04:04:00 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	293185	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	477119	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	470729	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	241462	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	178229	46.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.40%	
35) Dibromofluoromethane	5.47	113	140047	46.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.66%	
50) Toluene-d8	8.70	98	590677	50.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.44%	
62) 4-Bromofluorobenzene	11.12	95	233067	52.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.22%	
Target Compounds						
16) Acetone	2.42	43	4267	2.479	ug/l #	88
30) Chloroform	5.18	83	4763	0.811	ug/l	97
47) Bromodichloromethane	7.88	83	7264	1.509	ug/l	97
60) Dibromochloromethane	9.57	129	10070	2.713	ug/l	99
71) Bromoform	10.84	173	8788	2.961	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050720\
Data File : VX016095.D
Acq On : 07 May 2020 15:41
Operator : JC/SP
Sample : L2550-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z5-S1EFF-200430

Quant Time: May 08 04:04:00 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

