

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX050720\
 Data File : VX016096.D
 Acq On : 07 May 2020 16:04
 Operator : JC/SP
 Sample : L2551-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z4-EFF-200430

Quant Time: May 08 04:07:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 06 06:43:32 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	298308	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	488014	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	474955	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	247763	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	183859	46.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.68%	
35) Dibromofluoromethane	5.47	113	144636	46.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.56%	
50) Toluene-d8	8.70	98	596044	50.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.06%	
62) 4-Bromofluorobenzene	11.12	95	241742	53.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.70%	
Target Compounds						
16) Acetone	2.42	43	5441	3.106	ug/l	96
30) Chloroform	5.18	83	4695	0.786	ug/l	96
47) Bromodichloromethane	7.88	83	7113	1.445	ug/l	97
60) Dibromochloromethane	9.57	129	9994	2.633	ug/l	98
71) Bromoform	10.84	173	8758	2.924	ug/l #	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX050720\
Data File : VX016096.D
Acq On : 07 May 2020 16:04
Operator : JC/SP
Sample : L2551-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z4-EFF-200430

Quant Time: May 08 04:07:10 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X050420W.M
Quant Title : SW846 8260
QLast Update : Wed May 06 06:43:32 2020
Response via : Initial Calibration

