

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050721\
 Data File : VX021872.D
 Acq On : 07 May 2021 13:53
 Operator : JC/MD
 Sample : VX0507WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0507WBL01

Quant Time: May 08 06:19:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050721W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 08 05:38:18 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	77907	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	155809	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	137334	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	54355	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	58869	45.91	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.82%
35) Dibromofluoromethane	5.397	113	48857	45.76	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.52%
50) Toluene-d8	8.653	98	188643	46.30	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.60%
62) 4-Bromofluorobenzene	11.085	95	67464	41.95	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	83.90%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050721\
Data File : VX021872.D
Acq On : 07 May 2021 13:53
Operator : JC/MD
Sample : VX0507WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0507WBL01

Quant Time: May 08 06:19:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050721W.M
Quant Title : SW846 8260
QLast Update : Sat May 08 05:38:18 2021
Response via : Initial Calibration

