

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050721\
 Data File : VX021875.D
 Acq On : 07 May 2021 15:08
 Operator : JC/MD
 Sample : M2341-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: May 08 06:20:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050721W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 08 05:38:18 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	76413	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	152927	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	125460	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	45028	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	63664	50.63	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.26%
35) Dibromofluoromethane	5.397	113	51788	49.41	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.82%
50) Toluene-d8	8.653	98	198678	49.68	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.36%
62) 4-Bromofluorobenzene	11.085	95	67161	42.55	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.10%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050721\
Data File : VX021875.D
Acq On : 07 May 2021 15:08
Operator : JC/MD
Sample : M2341-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: May 08 06:20:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050721W.M
Quant Title : SW846 8260
QLast Update : Sat May 08 05:38:18 2021
Response via : Initial Calibration

