

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050721\
 Data File : VX021877.D
 Acq On : 07 May 2021 15:56
 Operator : JC/MD
 Sample : M2341-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: May 08 06:21:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050721W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 08 05:38:18 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	73905	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	147133	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	127412	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	51310	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	62040	51.01	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.02%	
35) Dibromofluoromethane	5.391	113	50924	50.50	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.00%	
50) Toluene-d8	8.653	98	192080	49.92	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.84%	
62) 4-Bromofluorobenzene	11.085	95	68729	45.26	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 90.52%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050721\
Data File : VX021877.D
Acq On : 07 May 2021 15:56
Operator : JC/MD
Sample : M2341-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: May 08 06:21:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050721W.M
Quant Title : SW846 8260
QLast Update : Sat May 08 05:38:18 2021
Response via : Initial Calibration

