

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050721\
 Data File : VX021871.D
 Acq On : 07 May 2021 13:03
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050721

Quant Time: May 08 06:02:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050721W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 08 05:38:18 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	104	0.00
2 T	Dichlorodifluoromethane	0.658	0.657	0.2	103	0.00
3 P	Chloromethane	0.791	0.758	4.2	101	0.00
4 C	Vinyl Chloride	0.913	0.879	3.7#	103	0.00
5 T	Bromomethane	0.333	0.323	3.0	109	0.00
6 T	Chloroethane	0.475	0.486	-2.3	102	0.00
7 T	Trichlorofluoromethane	1.258	1.202	4.5	101	0.00
8 T	Diethyl Ether	0.528	0.488	7.6	102	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.640	0.631	1.4	108	0.00
10 T	Methyl Iodide	0.760	0.758	0.3	98	0.00
11 T	Tert butyl alcohol	0.220	0.191	13.2	95	0.00
12 CM	1,1-Dichloroethene	0.678	0.628	7.4#	102	0.00
13 T	Acrolein	0.110	0.079	28.2#	97	0.00
14 T	Allyl chloride	1.242	1.170	5.8	101	0.00
15 T	Acrylonitrile	0.413	0.382	7.5	98	0.00
16 T	Acetone	0.369	0.370	-0.3	109	0.00
17 T	Carbon Disulfide	1.886	1.802	4.5	102	0.00
18 T	Methyl Acetate	1.023	0.926	9.5	98	0.00
19 T	Methyl tert-butyl Ether	2.174	2.096	3.6	102	0.00
20 T	Methylene Chloride	0.875	0.732	16.3	103	0.00
21 T	trans-1,2-Dichloroethene	0.723	0.692	4.3	101	0.00
22 T	Diisopropyl ether	2.313	2.208	4.5	101	0.00
23 T	Vinyl Acetate	2.276	2.205	3.1	100	0.00
24 P	1,1-Dichloroethane	1.330	1.292	2.9	104	0.00
25 T	2-Butanone	0.635	0.605	4.7	100	0.00
26 T	2,2-Dichloropropane	1.190	1.155	2.9	105	0.00
27 T	cis-1,2-Dichloroethene	0.854	0.823	3.6	103	0.00
28 T	Bromochloromethane	0.608	0.601	1.2	103	0.00
29 T	Tetrahydrofuran	0.377	0.349	7.4	96	0.00
30 C	Chloroform	1.356	1.276	5.9#	101	0.00
31 T	Cyclohexane	1.087	1.067	1.8	105	0.00
32 T	1,1,1-Trichloroethane	1.190	1.130	5.0	103	0.00
33 S	1,2-Dichloroethane-d4	0.823	0.794	3.5	103	-0.01
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	101	0.00
35 S	Dibromofluoromethane	0.343	0.341	0.6	104	0.00
36 T	1,1-Dichloropropene	0.530	0.525	0.9	105	0.00
37 T	Ethyl Acetate	0.640	0.618	3.4	99	0.00
38 T	Carbon Tetrachloride	0.530	0.525	0.9	102	0.00
39 T	Methylcyclohexane	0.595	0.600	-0.8	105	0.00
40 TM	Benzene	1.596	1.565	1.9	101	0.00
41 T	Methacrylonitrile	0.346	0.333	3.8	101	0.00
42 TM	1,2-Dichloroethane	0.547	0.536	2.0	100	0.00
43 T	Isopropyl Acetate	1.083	1.039	4.1	99	0.00
44 TM	Trichloroethene	0.388	0.388	0.0	102	0.00
45 C	1,2-Dichloropropane	0.418	0.412	1.4#	102	0.00
46 T	Dibromomethane	0.283	0.272	3.9	101	0.00
47 T	Bromodichloromethane	0.570	0.561	1.6	101	0.00
48 T	Methyl methacrylate	0.498	0.482	3.2	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050721\
 Data File : VX021871.D
 Acq On : 07 May 2021 13:03
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050721

Quant Time: May 08 06:02:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050721W.M
 Quant Title : SW846 8260
 QLast Update : Sat May 08 05:38:18 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.010	0.0	101	0.00
50 S	Toluene-d8	1.308	1.301	0.5	103	0.00
51 T	4-Methyl-2-Pentanone	0.674	0.647	4.0	97	0.00
52 CM	Toluene	1.019	0.993	2.6#	100	0.00
53 T	t-1,3-Dichloropropene	0.665	0.659	0.9	102	0.00
54 T	cis-1,3-Dichloropropene	0.696	0.694	0.3	102	0.00
55 T	1,1,2-Trichloroethane	0.407	0.402	1.2	101	0.00
56 T	Ethyl methacrylate	0.715	0.682	4.6	100	0.00
57 T	1,3-Dichloropropane	0.697	0.692	0.7	100	0.00
58 T	2-Chloroethyl Vinyl ether	0.350	0.339	3.1	97	0.00
59 T	2-Hexanone	0.519	0.512	1.3	99	0.00
60 T	Dibromochloromethane	0.439	0.420	4.3	100	0.00
61 T	1,2-Dibromoethane	0.425	0.421	0.9	101	0.00
62 S	4-Bromofluorobenzene	0.516	0.518	-0.4	104	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	101	0.00
64 T	Tetrachloroethene	0.322	0.326	-1.2	107	0.00
65 PM	Chlorobenzene	1.159	1.140	1.6	100	0.00
66 T	1,1,1,2-Tetrachloroethane	0.426	0.418	1.9	103	0.00
67 C	Ethyl Benzene	2.136	2.145	-0.4#	102	0.00
68 T	m/p-Xylenes	0.816	0.835	-2.3	103	0.00
69 T	o-Xylene	0.815	0.795	2.5	102	0.00
70 T	Styrene	1.355	1.369	-1.0	99	0.00
71 P	Bromoform	0.324	0.322	0.6	100	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
73 T	Isopropylbenzene	4.521	4.318	4.5	101	0.00
74 T	N-amyl acetate	2.423	2.256	6.9	99	0.00
75 P	1,1,2,2-Tetrachloroethane	1.677	1.528	8.9	100	0.00
76 T	1,2,3-Trichloropropane	1.422	1.344	5.5	98	0.00
77 T	Bromobenzene	0.974	0.916	6.0	99	0.00
78 T	n-propylbenzene	5.356	5.272	1.6	103	0.00
79 T	2-Chlorotoluene	3.245	3.134	3.4	103	0.00
80 T	1,3,5-Trimethylbenzene	3.885	3.816	1.8	102	0.00
81 T	trans-1,4-Dichloro-2-butene	0.613	0.600	2.1	102	0.00
82 T	4-Chlorotoluene	3.756	3.731	0.7	103	0.00
83 T	tert-Butylbenzene	3.766	3.673	2.5	103	0.00
84 T	1,2,4-Trimethylbenzene	3.897	3.878	0.5	103	0.00
85 T	sec-Butylbenzene	4.624	4.652	-0.6	104	0.00
86 T	p-Isopropyltoluene	3.941	3.958	-0.4	103	0.00
87 T	1,3-Dichlorobenzene	1.840	1.819	1.1	104	0.00
88 T	1,4-Dichlorobenzene	1.903	1.869	1.8	104	0.00
89 T	n-Butylbenzene	3.634	3.821	-5.1	105	0.00
90 T	Hexachloroethane	0.712	0.697	2.1	104	0.00
91 T	1,2-Dichlorobenzene	1.816	1.780	2.0	102	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.404	0.344	14.9	99	0.00
93 T	1,2,4-Trichlorobenzene	1.094	1.084	0.9	105	0.00
94 T	Hexachlorobutadiene	0.427	0.426	0.2	104	0.00
95 T	Naphthalene	4.482	4.371	2.5	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050721\
Data File : VX021871.D
Acq On : 07 May 2021 13:03
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050721

Quant Time: May 08 06:02:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050721W.M
Quant Title : SW846 8260
QLast Update : Sat May 08 05:38:18 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.109	1.084	2.3	103	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6