

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX050724\  
 Data File : VX041396.D  
 Acq On : 07 May 2024 10:12  
 Operator : JC/MD  
 Sample : VX0507WBL01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0507WBL01

Quant Time: May 08 01:55:53 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X043024W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed May 01 02:38:17 2024  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|-------|----------|
| -----                       |        |       |          |          |       |          |
| Internal Standards          |        |       |          |          |       |          |
| 1) Pentafluorobenzene       | 5.550  | 168   | 174583   | 50.000   | ug/l  | # 0.00   |
| 34) 1,4-Difluorobenzene     | 6.763  | 114   | 250153   | 50.000   | ug/l  | 0.00     |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 242972   | 50.000   | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152   | 117602   | 50.000   | ug/l  | 0.00     |
| System Monitoring Compounds |        |       |          |          |       |          |
| 33) 1,2-Dichloroethane-d4   | 5.952  | 65    | 91333    | 48.055   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | =     | 96.120%  |
| 35) Dibromofluoromethane    | 5.385  | 113   | 76873    | 45.347   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =     | 90.700%  |
| 50) Toluene-d8              | 8.647  | 98    | 246175   | 45.881   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | =     | 91.760%# |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 106424   | 51.239   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =     | 102.480% |

|                  |        |
|------------------|--------|
| Target Compounds | Qvalue |
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX050724\  
Data File : VX041396.D  
Acq On : 07 May 2024 10:12  
Operator : JC/MD  
Sample : VX0507WBL01  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0507WBL01

Quant Time: May 08 01:55:53 2024  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X043024W.M  
Quant Title : SW846 8260  
QLast Update : Wed May 01 02:38:17 2024  
Response via : Initial Calibration

