

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050724\
 Data File : VX041403.D
 Acq On : 07 May 2024 13:54
 Operator : JC/MD
 Sample : P2389-03 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1402

Quant Time: May 08 02:00:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 01 02:38:17 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	186538	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	229996	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	215832	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	122225	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	94666	46.617	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.240%
35) Dibromofluoromethane	5.385	113	79343	50.906	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.820%
50) Toluene-d8	8.647	98	262117	53.134	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	106.260%
62) 4-Bromofluorobenzene	11.079	95	93650	49.040	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.080%
Target Compounds						
16) Acetone	2.392	43	1099408	1244.975	ug/l	# Qvalue 1

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050724\
Data File : VX041403.D
Acq On : 07 May 2024 13:54
Operator : JC/MD
Sample : P2389-03 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1402

Quant Time: May 08 02:00:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 01 02:38:17 2024
Response via : Initial Calibration

