

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050724\
 Data File : VX041408.D
 Acq On : 07 May 2024 15:50
 Operator : JC/MD
 Sample : P2389-01 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1400

Quant Time: May 08 02:02:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 01 02:38:17 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	177849	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	261458	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	205168	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	106384	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	92595	47.825	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.640%
35) Dibromofluoromethane	5.385	113	76051	42.922	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	85.840%
50) Toluene-d8	8.647	98	234313	41.782	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	83.560%#
62) 4-Bromofluorobenzene	11.079	95	87163	40.151	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	80.300%#
Target Compounds						
					Qvalue	
16) Acetone	2.398	43	89734	102.946	ug/l	# 81
20) Methylene Chloride	2.788	84	2355	1.401	ug/l	# 88
40) Benzene	6.037	78	2560	0.392	ug/l	# 88
52) Toluene	8.720	92	1445	0.362	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050724\
Data File : VX041408.D
Acq On : 07 May 2024 15:50
Operator : JC/MD
Sample : P2389-01 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1400

Quant Time: May 08 02:02:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 01 02:38:17 2024
Response via : Initial Calibration

