

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050724\
 Data File : VX041410.D
 Acq On : 07 May 2024 16:36
 Operator : JC/MD
 Sample : P2410-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 A2295

Quant Time: May 08 02:02:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 01 02:38:17 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	195897	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	288232	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	269216	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	136545	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	105132	49.297	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.600%
35) Dibromofluoromethane	5.385	113	86895	44.487	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.980%
50) Toluene-d8	8.647	98	315955	51.107	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.220%
62) 4-Bromofluorobenzene	11.079	95	119914	50.106	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.220%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	38391	37.556	ug/l	94
25) 2-Butanone	4.581	43	18702	13.549	ug/l	94
29) Tetrahydrofuran	5.026	42	10753	11.912	ug/l	100
52) Toluene	8.720	92	1463	0.332	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050724\
Data File : VX041410.D
Acq On : 07 May 2024 16:36
Operator : JC/MD
Sample : P2410-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
A2295

Quant Time: May 08 02:02:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X043024W.M
Quant Title : SW846 8260
QLast Update : Wed May 01 02:38:17 2024
Response via : Initial Calibration

